



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R21/2007-NL1-12.02
Project number 12200118
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Microtek S.r.l. Via Des Giavis 33010 Pagnacco - UD Italy
Manufacturer	Microtek S.r.l. Via Des Giavis 33010 Pagnacco - UD Italy
Identification of the certified type	An electronic taximeter Type : MT9; MT12; MTMK
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R21 - Edition 2007

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
24 October 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R21/2007-NL1-12.02
Project number 12200118
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

Number NMI-10201071-01 dated 27 February 2012 that includes 44 pages;
Number NMI-10201071-02 dated 27 February 2012 that includes 11 pages;
Number NMI-12200118-01 dated 24 October 2012 that includes 21 pages.

Characteristics of the electronic taximeter:

Power Supply	: 9 - 16 V DC
Range of device constant	: $k = 500 \text{ km}^{-1}$ to 7500 km^{-1} ;
Resolution of device constant	: 1 km^{-1} ;
Range of time tariff	: 0,00 ¹ CU/h to 1800,00 CU/h;
Resolution of time tariff	: 0,05 CU/h;
Range of distance tariff	: 0,00 CU/km to 500,00 CU/km;
Resolution of distance tariff	: 0,05 CU/km;
Time measuring signal frequency	: 10 Hz
Temperature range	: -10 °C / +70 °C

¹ CU = Currency Units