

Member State
P.R.China

OIML Certificate No.
R31/1995-CN1-11.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 P.R.China

Applicant

Name: Dandong Gas Meter Co., Ltd.
Address: No.66 Gucheng Road, Yuanbao District, Dandong, Liaoning, (118003) P.R. China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:
Address:

Identification of the certified type

Diaphragm gas meter

Type	J2.5C II
Accuracy class	1.5%
Maximum flow rate	4 m ³ /h
Minimum flow rate	0.025 m ³ /h
Cyclic volume of a gas volume meter	1.2 dm ³
Maximum operating pressure	50 kPa
Operating temperature range	(-10~+40) °C

The numbers of the associated test reports and the included pages:
20100M001 14 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 31

Edition: 1995

For accuracy class (if applicable): 1.5%

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority
Office of OIML Affairs

中国秘书处

June 16, 2011

The OIML member

蒲长城

June 16, 2011

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.