

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Elektrometal S.A.
Address: ul. Stawowa 71
43-400 Cieszyn
Poland

Manufacturer of the certified type

Name: Elektrometal S.A.
Address: ul. Stawowa 71
43-400 Cieszyn
Poland

Identification of the certified type

Diaphragm gas meter, with the following characteristics:

The meter, without physical changes, can be classified into one of the following G-values:

model	Q_{\max} (m ³ /h)	Q_{\min} (dm ³ /h)	p_{\max} (bar)	V (dm ³)
EM-G4	6	40	0,5	1,2
EM-G2,5	4	25	0,5	1,2
EM-G1,6	2,5	16	0,5	1,2

OIML Member State
The Netherlands

This Certificate attests the conformity of the above identified type (represented by the samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R31

Edition 1995 (E)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated Test Report: N° R31/1995-NL1-07.03 that includes 15 pages.

The design of the above mentioned gas meters is identical to the type of the gas meters as described in the OIML Certificate of Conformity N° R31/1995-NL1-07.03, manufactured by the company i-meter. The manufacturer produces the diaphragm gas meters based on a license agreement with the company i-meter, for whom the above mentioned OIML Certificate of Conformity was issued previously.

The Issuing Authority NL1
NMI Certip, 22 July 2009



C. Oosterman
Head Certification Board

*
**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.