

Member State
P.R.China

OIML Certificate No.
R46/2012-CN1-17.01



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 P.R.China

Applicant

Name: Zhejiang CHINT Instrument & Meters Co., Ltd

Address: Wenzhou Bridge Industrial Zone, Yueqing, Zhejiang, (325603) P.R. China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Electricity Meter

Type	DDZY666-Z
Accuracy class	B
Nominal voltage /Nominal frequency	220V/50Hz
Minimum current/ Transitional current/ Maximum current	0.25-0.5(60)A
Connection modes	single-phase, two-wire, 1 element
Energy flow direction /Register	bi-directional / two-register
Indoor and outdoor type	Indoor
Insulation protection class	Category 2
Temperature limit	-40°C — + 70°C
Humidity class	H2
Hardware version/Software version	ZTY8.067.1601/0118201410232301

The numbers of the associated test reports and the included pages:

ZJIM-D201607001 44 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 46-1/-2 (2012) Active Electrical Energy Meters

Edition: 2012

For accuracy class (if applicable): B

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports.

This Revision replaces the preceding certificate.

The issuing authority


Office of OIML Affairs

March 20, 2017

The CIML member



March 20, 2017

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.