



OIML Certificate

OIML Member StateThe Netherlands



Number R46/2012-A-NL1-21.06 revision 1 Project number 2534036 Page 1 of 3

Issuing authority NMi Certin B.V.

Person responsible: M. Boudewijns

+

Applicant and Manufacturer

Ningbo Sanxing Medical & Electric Co. Ltd.

No.1166 Mingguang North Road

Yinzhou Industrial Zone, Jiangshan Town, Yinzhou District

Ningbo City, Zhejiang Province

China

Identification of the certified type

A measuring instrument

Type: S34U18 (160A)

Characteristics

See page 2 and further

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 46-1/-2 (2012) "Active electrical energy meters"

Accuracy class B

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority NMi Certin B.V., OIML Issuing Authority NL1

4 March 2021

Certification Board







This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.









OIML Certificate





Number R46/2012-A-NL1-21.06 revision 1 Project number 2534036 Page 2 of 3



The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMi-2534036-01 dated 14 January 2021 that includes 21 pages;
- No. NMi-2534036-02 dated 14 January 2021 that includes 12 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented.

Table 1 General characteristics

General characteristics	
Meter type	Static
Connection mode (phase, wires, elements)	3p, 4w, 3e
Direction of energy flow / registers	Two-registers, bi-directional
Terminal arrangement	DIN
Protective class	Category 2
Environmental application	
Ambient temperature range	-10 °C to +70 °C* *tested for the upper temperature of +75 °C on customer request.
Humidity class	H1
IP Rating / environmental use	IP54
Meter quantities	
Nominal voltage (U _{nom})	3x133/230V3x230/400V
Nominal frequency (f _{nom})	60 Hz
Maximum current (/ _{max})	160 A
Transitional current (Itr)	2 A (I _b = 20 A)
Minimum current (I _{min})	1 A
Starting current (I _{st})	0,08 A
Meter constant	1.000 imp./kWh
Product version	
Hardware version	S34U18 S09.Y2.J2 M20
Software identification	P4281 2020-08-24 CHECKSUM: F9EA





OIML Certificate



OIML Member State The Netherlands

Number R46/2012-A-NL1-21.06 revision 1 Project number 2534036 Page 3 of 3



Revision History

This revision replaces the previous versions.

Revision	Date	Change(s)
Initial	19-01-2021	Initial issue.
1	04-03-2021	Editorial changes: Additional remark added regarding the temperature range.









