

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate N°
R49-1/2003-DE1-07.02

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Gudrun Wendt

Applicant

Name: Sensus Metering Systems GmbH Ludwigshafen
Address: Industriestr. 16, 67063 Ludwigshafen am Rhein

Manufacturer of the certified type is the applicant.

Identification of the certified type

Water meter intended for the metering of cold potable Water
(mechanical, complete)
Type: Model 620

Further characteristics see page 3

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R49-1 (2003): Metrological and technical requirements
R49-2 (2004): Test methods
R49-3 (2004): Test report format

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R49-1/2003-DE1-07.02

The conformity was established by the results of tests and examinations provided in the associated Report No. PTB-1.5-4025660-a and Test Report No. PTB-1.5-4025660-b.

The Issuing Authority

The OIML Member

Dr. Gudrun Wendt

Dr. Roman Schwartz

Head of Department
Liquid Flow

Head of Division
Mechanics and Acoustics

17.04.2007

17.04.2007

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R49-1/2003-DE1-07.02

Identification of the certified pattern – page 1 continued

Metrological characteristics:

Q_1 -	0,0127 m ³ /h	0,00794 m ³ /h
Q_2 -	0,0203 m ³ /h	0,0127 m ³ /h
Q_3 -	4 m ³ /h	2,5 m ³ /h
Q_4 -	5 m ³ /h	3,125 m ³ /h
Q_2/Q_1 -	1,6	1,6
Q_3/Q_1 -	315	315
length	190mm	170mm
thread	G 1 B	G ¾ B

Measuring principle:	Rotary Piston
Accuracy Class:	2
Environmental class:	C (M2)
Maximum admissible temperature:	30 °C
Maximum admissible pressure:	1,6 MPa (16 bar)

Installation details:

Connection type:	screw thread
Minimum straight length of inlet pipe:	0 mm
Minimum straight length of outlet pipe:	0 mm
Flow conditioner:	not required
Orientation:	horizontal or vertical