



Member State
Switzerland

OIML Certificate
No R49/2006-CH1-10.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name Federal Office of Metrology METAS
Certification Body METAS-Cert

Address METAS, Lindenweg 50, CH-3003 Bern-Wabern

Person responsible Jürg Ramseyer, Head of METAS-Cert

Applicant

Name E.WEHRLE GmbH

Address Obertalstrasse 8, D – 78120 Furtwangen

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified pattern

**Single-jet impeller meter with wet register intended for the
metering of cold water (T30)**

Type

**ENK-EA., ENK-EAV, ENK-EAN, ENK-EAO
ESK-EA., ESK-EAV, ESK-EAN, ESK-EAO**

For further characteristics see page 3 and ff.

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 49-1, edition 2006
for accuracy class 2

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



**OIML Certificate
No R49/2006-CH1-10.01**

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

No 135-11181 that includes 4 pages

The Issuing Authority

Jürg Ramseyer, Head of METAS-Cert

The OIML Member

Dr. Philippe Richard, Vice Director

CH-3003 Bern-Wabern, March 3, 2010

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and the associated Test Report is not permitted, although either may be reproduced in full.



OIML Certificate No R49/2006-CH1-10.01

1 Description of the type

The meter is a single-jet wet-rotor meter with various threading. The measuring sensor and the register are tightly interconnected to prevent any unauthorized opening and manipulation of the meter

2 Technical specifications

Q ₃	m ³ /h	2,5	4,0
Q ₄	m ³ /h	3,125	5,0
Q ₂ /Q ₁		1,6	1,6
Overall length	mm	≥ 110	≥ 115
Nominal Diameter	DN	15	20
Threaded connector of the body		≥ G ¾ B	≥ G 1 B
Q ₁ mounting horizontal	ℓ/h	15,6 / 20 / 25 / 31,3 / 39,7 / 50 / 62,5 / 79,4 / 100 / 125 / 156,3 / 200 / 250	25 / 32 / 40 / 50 / 63 / 80 / 100 / 127 / 160 / 200 / 250 / 320 / 400
Q ₁ mounting vertical	ℓ/h	31,3 / 39,7 / 50 / 62,5 / 79,4 / 100 / 125 / 156,3 / 200 / 250	50 / 63 / 80 / 100 / 127 / 160 / 200 / 250 / 320 / 400
Q ₂ mounting horizontal	ℓ/h	25 / 32 / 40 / 50 / 63,5 / 80 / 100 / 127 / 160 / 200 / 250 / 320 / 400	40 / 51,2 / 64 / 80 / 100,8 / 128 / 160 / 203,2 / 256 / 320 / 400 / 512 / 640
Q ₂ mounting vertical	ℓ/h	50 / 63,5 / 80 / 100 / 127 / 160 / 200 / 250 / 320 / 400	80 / 100,8 / 128 / 160 / 203,2 / 256 / 320 / 400 / 512 / 640
Measuring range (Q ₃ /Q ₁), mounting horizontal		160 / 125 / 100 / 80 / 63 / 50 / 40 / 31,5 / 25 / 20 / 16 / 12,5 / 10	160 / 125 / 100 / 80 / 63 / 50 / 40 / 31,5 / 25 / 20 / 16 / 12,5 / 10
Measuring range (Q ₃ /Q ₁), mounting vertical		80 / 63 / 50 / 40 / 31,5 / 25 / 20 / 16 / 12,5 / 10	80 / 63 / 50 / 40 / 31,5 / 25 / 20 / 16 / 12,5 / 10
Pressure loss class ΔP		63	63
Calibration value	ℓ	0,05	
Water pressure MAP	bar	16	
Temperature class		T30: 0,1°C ≤ T ≤ 30°C	



OIML Certificate
No R49/2006-CH1-10.01

Accuracy class		$\pm 2 \% (Q_2 \leq Q \leq Q_4)$
		$\pm 5 \% (Q_1 \leq Q \leq Q_2)$
Environmental classification		Class B 5 °C up to 55 °C
Flow profile sensitivity class		U0 / D0