

# Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML  
Germany



OIML Certificate N°  
**R51/1996-DE1-05.05**

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. Roman Schwartz

### Applicant

Name: Weber-Waagenbau und Wägeelektronik GmbH  
Address: Boschstr. 5-7  
68753 Waghäusel  
Germany

Manufacturer of the certified type is the applicant.

**Identification of the certified type** Automatic catchweighing instrument  
Type: CW55

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R 51-1**, edition 1996  
for accuracy class X(1)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

# Physikalisch-Technische Bundesanstalt

OIML Certificate N°  
**R51/1996-DE1-05.05**

The conformity was established by the results of tests and examinations provided in the associated Report

No. 1.12-4016368 (12 pages)

and the associated Test Reports

No. 1.12-4016368/1 (58 pages) and

No. 1.12-4016368/2 (16 pages).

## The Issuing Authority

Dr. R. Schwartz  
Direktor und Professor

2005-12-05

## The OIML Member

Prof. Dr. M. Kochsiek  
Vizepräsident

2005-12-05

## Identification of the type (continued)

Automatic electromechanical weighing instrument, working as checkweigher without lever work, performed as single or multiple range as well as single or multi interval instrument.

Accuracy class	X(1)
Mode of operation	Start-stop operation
Number of weighings per minute	$\leq 24$
Power supply voltage	230 V / 400 V AC, 50/60 Hz
Temperature range	-10 °C / +40 °C
Verification scale interval e	$\geq 1$ g
Number n of verification scale intervals	$\leq 4000$
Maximum load Max	$\leq 400$ kg
Minimum load Min	$\geq 20 e_1$
Number of intervals or ranges	$\leq 3$
Ratio between $e_1$ und $e_{i+1}$	$\leq 3$
Additive tare	$\leq -10\% \cdot \text{Max}$
Subtractive tare	$\leq 100\% \cdot \text{Max}$

**Table 3.1: Technical data of the weighing instrument of type CW55**

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.