

	 Physikalisch-Technische Bundesanstalt Braunschweig und Berlin
OIML Member State Germany	OIML Certificate No. R60/2000-A-DE1-2019.03
OIML CERTIFICATE ISSUED UNDER SCHEME A	
OIML Issuing Authority Name: Physikalisch-Technische Bundesanstalt, Conformity Assessment Body Address: Bundesallee 100, 38116 Braunschweig, GERMANY Person responsible: Hon.-Prof. Dr. R. Schwartz	
Applicant Name: Hottinger Baldwin Messtechnik GmbH Address: Im Tiefen See 45, 64293 Darmstadt, Germany	
Manufacturer Name: Hottinger Baldwin Messtechnik GmbH Address: Im Tiefen See 45, 64293 Darmstadt, Germany	
Identification of the certified type <i>(the detailed characteristics will be defined in the additional pages)</i> Type: Z6R	
Designation of the module <i>(if applicable)</i> Load cell	
This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML): OIML R 60 Edition (year): 2000 For accuracy class (if applicable): C3	

OIML Certificate No.
R60/2000-A-DE1-2019.03

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No. 1.12-4093446 dated 03.07.2019 that includes 6 pages

The technical documentation relating to the identified type is contained in documentation file:

No. ZDS-R60/2000-A-DE1-2018.03 dated 03.07.2019 that includes 2 pages


The Certificate and Test-Reports are based of the OIML-MAA-Certificate R60/2000-DE1-15.01 and have been transferred due to Procedural Documents OIML-CS PD-07.

OIML Certificate History

Revision No.	Date	Description of the modification
First issuance	03.07.2019	---

Identification, signature and stamp

The Issuing Authority



Dipl.-Ing. K. Schulz

Member of Conformity Assessment Body



The OIML Member



Hon.-Prof. Dr. R. Schwartz

Vice President of PTB

Date: 03.07.2019

Table 1: Essential data

Accuracy class		C3	D1
Maximum number of load cell intervals n_{LC}		3000	1000
Rated output	mV/V	2	
Maximum capacity E_{max}	kg	20 / 22 / 30 / 33 / 50 / 55 / 100 / 110 / 200 / 220	
Minimum load cell verification interval $v_{min} = (E_{max} / Y)$		$E_{max} / 11111$	

Minimum dead load: $0\% \cdot E_{max}$; Safe overload: $150\% \cdot E_{max}$; Input impedance: 350Ω

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML
Certification System