

**OIML Member State**United Kingdom of Great Britain
and Northern Ireland**OIML Certificate No****Annex 01 to R60/2000-GB1-05.06**
Revision 2**ANNEX to an OIML Certificate of Conformity**OIML Issuing Authority **NMO**
Stanton Avenue
Teddington
TW11 0JZ
United KingdomPerson responsible: **Mannie Panesar – Head of Technical Services**Applicant **Societa' Cooperativa Bilanciai Srl**
Via S. Ferrari, 16
41011 Campogalliano
Campogalliano (Modena)
ItalyManufacturer **The applicant****Correction:**

In the Characteristics of the Load Cell table (Page 2):

The value of Relative DR, Z = 3000 is corrected to Z = 6000

<i>Model designation</i>	<i>Designation</i>	<i>Value</i>						<i>Units</i>
Classification		C1	C2	C3	C4	C5	C6	
Relative DR (ratio to minimum dead load output return)	$Z = E_{max}/(2*DR)$			6000				

The OIML Issuing Authority

Issue date: 14 September 2018

Grégory Glas**Lead Technical Manager***For and on behalf of the Head of Technical Services*