



OIML Certificate

OIML Member State
The Netherlands

Number R60/2000-A-NL1-18.33
Project number 1902131
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and Manufacturer Mettler-Toledo (Changzhou) Precision Instruments Ltd.
No. 22, Zhengqiang Road, XinBei District
Changzhou, Jiangsu 213125
Peoples Republic of China

Identification of the certified type A **single point load cell**, with strain gauges.
Type : MT1260 series

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60 - Edition 2000 (E) for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
20 December 2018


C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org





OIML Certificate

OIML Member State
The Netherlands

Number R60/2000-A-NL1-18.33
Project number 1902131
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R60/2000-NL1-03.14A dated 7 October 1998 that includes 37 pages;
- No. R60/2000-NL1-03.14B dated 4 July 2003 that includes 40 pages;
- No. NMI-1902131-03 dated 8 November 2018 that includes 16 pages;
- No. NMI-1902131-04 dated 8 November 2018 that includes 15 pages.

Characteristics of the load cell:

Maximum capacity (E_{\max})	50 kg up to and including 750 kg
Minimum dead load	0 kg
Accuracy Class	C
Rated Output	2,0 mV/V \pm 0,2 mV/V
Maximum number of load cell intervals (n) ⁽¹⁾	3000
Ratio of minimum LC Verification interval ⁽¹⁾ $Y = E_{\max} / v_{\min}$	10000
Ratio of minimum dead load output return ⁽¹⁾ $Z = E_{\max} / (2 * DR)$	3000
Input impedance	410 $\Omega \pm 10 \Omega$
Temperature range	-10 °C / +40 °C
Fraction p_{LC}	0,7
Humidity Class	CH
Safe overload	150 % of E_{\max}
Output impedance	350 $\Omega \pm 4 \Omega$
Recommended excitation	5 - 15 V AC / DC
Excitation maximum	20 V AC / DC
Transducer material	Aluminium alloy
Atmospheric protection	Silicon rubber

Remarks:

1. The characteristics for n_{\max} , Y and Z can be reduced separately.

Each load cell produced is provided with an accompanying document with information about its characteristics.