



Australian Government
 Department of Industry,
 Innovation and Science



OIML Member State
 Australia

**National
 Measurement
 Institute**

OIML Certificate No.
R76-1/1992-AU1-16.01
 NMI file: R2015/218

OIML BASIC CERTIFICATE OF CONFORMITY

Issuing Authority National Measurement Institute
 Bradfield Rd
 Lindfield NSW 2070
 Australia

**Applicant and
 Manufacturer** Conweigh Innovation Pty Ltd
 1/58 Frederick Street
 Northgate QLD 4013
 Australia

**Identification of the
 certified pattern:** A Conweigh model CONW non-automatic weighing instrument.
 Further characteristics on page 2.

This Certificate attests the conformity of the above identified pattern (represented by the sample or samples identified in the OIML Basic Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 1992 for accuracy class

This Certificate relates only to the metrological and technical characteristics of the pattern of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Certificate history:

Issue no.	Date	Description of the modification
1	29 Aug 2016	Initial issue

The OIML Issuing Authority

Dr Amanda Rawlinson
 Manager, Pattern Approval
 Date: 29/8/16

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML Basic Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

**OIML Member State
Australia**

**OIML Certificate No.
R76-1/1992-AU1-16.01
NMI file: R2015/218**

The conformity was established by the results of tests and examinations provided in the associated Test Report:

NMI Australia report no. R2015/218[1],
dated 22nd August 2016, that includes 40 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	
Maximum capacity	Max = 40 000 kg
Verification scale interval	e = 50 kg
Weighing range(s)	Single interval
Maximum number of scale intervals	n <= 800 divisions
Temperature range	-10 °C / +40 °C
Power supply voltage	Internal rechargeable battery
Application	Weighing of shipping containers and similar
Software identification	Tested with Conweigh Tech App version 6.1.1

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML Basic Type Evaluation Report(s) is not permitted, although either may be reproduced in full.