



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-17.20
Project number 1900004
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and Manufacturer GRUPO EPELSA, S.L.
c/. Punto Net, 3
Polígono Industrial TECNOALCALÁ
E-28805 Alcalá de Henares
Madrid
Spain

Identification of the certified type **A Non-automatic weighing instrument**
Type : MC-40 PLUS

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)** **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
18 April 2017


C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-NL1-17.20
Project number 1900004
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-13200261-01 dated 16 September 2013 that includes 49 pages;
- No. R60/1991-NL-98.04 that includes 38 pages;
- No. R60/1991-NL-99.11A that includes 36 pages;
- No. R60/1991-NL-99.11B that includes 36 pages;
- No. R60/2000-NL-01.16 that includes 39 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III and IIII
Maximum number of verification scale intervals	$n \leq 7500$ for class III instruments $n \leq 1000$ for class IIII instruments
Temperature range	-10 °C / +40 °C
Weighing range(s)	Single interval Multi-interval Multiple range
Power supply voltage	100 – 240 V AC 50/60 Hz 24 V DC
Maximum number of load platforms	2
Software identification	XXXX.X.XXX.0X – X may be any number between 0 to 9 or a character between A to Z. The '0' indicates the metrological version of the software.