



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-17.22
Project number 16200726
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Kingship Weighing Machine Corp. No. 739, Renhua Road, Dali Dist., Taichung City 412 Taiwan, R.O.C.
Identification of the certified type	A Non-automatic weighing instrument Type : Z-6910MS
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
24 April 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-NL1-17.22
Project number 16200726
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-16200726-01 dated 19 April 2017 that includes 60 pages;
- No. NMI-16200726-02 dated 19 April 2017 that includes 33 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$0 \leq \text{Max} \leq 15 \text{ kg}$
Verification scale interval	$e \geq 2 \text{ g}$
Weighing range(s)	Multiple range
Maximum number of scale intervals (multiple range)	$n \leq 3000$ divisions (per weighing range)
Maximum number of weighing ranges	2
Tare	$T \leq \text{Max}_1 - e_1$ for multiple range instruments
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	100 – 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public
Software identification	Version number: 2.00