

OIML Member State

The Netherlands

OIML Certificate

Number R117/2019-A-NL1-24.07 revision 1 Project number 3927432 Page 1 of 4

Issuing authority NMi Certin B.V. Person responsible: M.Ph.D. Schmidt Applicant and Micro Motion, Inc.

Applicant and Micro Motion, Inc. Manufacturer 7070 Winchester Circle Boulder, Colorado 80301 United States of America

Identification of the certified type A **measurement device** (ultrasonic sensor with belonging electronics) intended to be used as a part of an interruptible or non-interruptible dynamic measuring system for liquids other than water.

Manufacturer's mark: Type: Rosemount Model 3814 and 3818

Characteristics See following page(s)

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 117-1: 2019 "Dynamic measuring systems for liquids other than water"

Accuracy class 0,3

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Reports is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V.

Thijsseweg 11

2629 JA Delft

certin@nmi.nl

www.nmi.nl

the Netherlands

T +31 88 636 2332

NMi Certin B.V., OIML Issuing Authority NL1 10 March 2025

Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.









OIML Certificate

Number R117/2019-A-NL1-24.07 revision 1 Project number 3927432 Page 2 of 4

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMi-3831466-01 dated 8 November 2024 that includes 55 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented. In Table 2 the characteristics of the family of instruments are presented. The construction of the measuring instrument is recorded in the Documentation folder no. TC8224-5.

Table 1 General characteristics

4
430 kHz - 1,0 MHz
60 °
See table 2; Bi-directional (if calibration performed in both flow directions)
See table 2
300 bar(g)
M2 / E2
-40 °C / +55 °C; H3, Condensing Open location
See table 3
liquid petroleum and related products, liquids food and chemical products in liquid state
10,4 36 VDC
See table 4





 (\bullet)

Number R117/2019-A-NL1-24.07 revision 1 Project number 3927432 Page 3 of 4

Table 2 General characteristics of the family of instruments

Meter size	Minimum flow rate [m³/h]	Maximum flow rate [m³/h]	Minimum Reynolds number [-]	Minimum measured quantity [m³]
4″	36	360	10000	1
6″	82	720	10000	2
8″	142	1450	6000	10
10"	223	2250	6000	20
12"	300	3170	6000	50
16″	300	4890	10000	100
18″	450	6334	10000	100
20" [1]	400	7880	10000	200
24″ [1]	580	11390	10000	200

^[1] Meter sizes that can be used only with the LT-05, LT-07 and LT-09 type transducers.

Table 3 General characteristics of the transducers

4"-10" Meter Sizes	12" - 24" Meter Sizes	Liquid Temperature Range	Viscosity Range
LT-01	LT-03	-50 °C to 100 °C	Up to 130 cst
LT-04	LT-05	-50 °C to 150 °C	Up to 1000 cst
LT-06	LT-07	-200 °C to 60 °C	Up to 130 cst
LT-08	LT-09	-50 °C to 150 °C	Up to 130 cst

Table 4 Software specifications

Part	Version	Checksum	
Operating system, Kernel	2.6.37.6	0babf2ffe2512c50c22b9af006dd97a7 ulmage.zip	
File system	1.02	feef7a9450ac3fe5cebccac5badd6894 rootfs.zip	
Firmware	1.04	a3800fb0ed46527d08371c141315b6b1 firmware.zip	Ċ





OIML Certificate

Number R117/2019-A-NL1-24.07 revision 1 Project number 3927432 Page 4 of 4

	Part	Version	Checksum
	Program checksum;	1.13	0130745426
)	Combined kernel, File System &	1.22	3499386616
	Firmware	1.24	1869761847
		1.27	1547356617
		1.42	0938002774
		1.60	4058185302
		1.61	3742344397
		1.70	2317473693

Certificate history:

Revision	Date	Description of the modification
0	8 November 2024	Initial release
1	10 March 2025	Addition of software version 1.70.
	(🗕)	