



OIML Member State

Italy

OIML Certificate No.

R137/2012-A-IT2-26.02

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority

Name: Tifernogas S.r.l.
Address: Via R. Morandi 44/d, 06012 Città di Castello (PG), Italy
Person responsible: Paolo Tafani Alunno

Applicant

Name: **Mesura Metering S.r.l.**
Address: **Via Statale , 11/13, Fraz. Ponte S. Marco
25011 Calcinato (BS), Italy**

Manufacturer

Name: Mesura Metering S.r.l.
Address: Via Statale , 11/13, Fraz. Ponte S. Marco
25011 Calcinato (BS), Italy

Identification of the certified type (the detailed characteristics will be defined in the additional pages)

Type: Ultrasonic gas meter
PRODIGI 2.0 EXT-EU
PRODIGI 2.0 40-ES

Designation of the module (if applicable)

Not applicable

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 137

Edition (year): 2012, including Amendment 2014

For accuracy class (if applicable): 1.5

OIML Certificate No. R137/2012-A-IT2-26.02																	
This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above. This OIML Certificate does not bestow any form of legal international approval.																	
The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report: No. TER26.03 dated 24 July 2026 that includes 43 pages																	
The technical documentation relating to the identified type is contained in documentation repository: No. QW 2024/000081																	
OIML Certificate History <table border="1"> <thead> <tr> <th>Revision No.</th> <th>Date</th> <th>Description of the modification</th> </tr> </thead> <tbody> <tr> <td>0</td> <td>24/07/2026</td> <td>Certificate Issuing</td> </tr> <tr> <td> </td> <td> </td> <td> </td> </tr> <tr> <td> </td> <td> </td> <td> </td> </tr> <tr> <td> </td> <td> </td> <td> </td> </tr> </tbody> </table>			Revision No.	Date	Description of the modification	0	24/07/2026	Certificate Issuing									
Revision No.	Date	Description of the modification															
0	24/07/2026	Certificate Issuing															
The OIML Issuing Authority Paolo Tafani Alunno Head of OIML Issuing Authority Date: 24 July 2026																	
<i>Important note:</i> Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.																	


General information of the instrument type

Manufacturer's trade mark	 A Cavagna Group Company
Type designation	PRODIGI 2.0 EXT-EU ; PRODIGI 2.0 40-ES
Accuracy class	1,5
Connection threads size	1"1/4 (PRODIGI 2.0 EXT-EU) ; 7/8" (PRODIGI 2.0 40-ES)
Environmental classes	M2 / E1
Ambient temperature range	-25°C to +55°C
Gas temperature range	-25°C to +55°C
Base pressure	Not applicable
Base temperature	15°C (only for PRODIGI 2.0 EXT-EU)
t_{sp}	15°C (only for PRODIGI 2.0 EXT-EU)
Orientation	Connection Ports vertical
Flow direction	Forward flow only
Indicating device	Electronic
Ancillary devices	NB-IoT or LTE-M or GPRS communication
Electrical power	Lithium battery, 3.6V
Software identification	11.3 – CRC: DFEDCF00


Additional information of the instrument type

Designation	Q_{max} [m ³ /h]	Q_t [m ³ /h]	Q_{min} [m ³ /h]	Cases material	P_{max} [bar]	Conn. Dist. [mm]
PRODIGI 2.0 EXT-EU	6	0.6	0.016	Steel	0.5	110
PRODIGI 2.0 40-ES	6	0.6	0.04	Steel	0.5	160

Labels




PRODIGI 2.0 EXT-EU marking



PRODIGI 2.0 40-ES marking


Sealing



Where:

1 – 2: Metrological seals

A – B: Screws for securing the transparent cover (non-metrological)



Metrological seal