



OIML Certificate

OIML Member State
The Netherlands

Number R137/2012-A-NL1-23.04 revision 1
Project number 3711995
Page 1 of 5

Issuing authority NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer RMG Messtechnik GmbH
Otto-Hahn-Strasse 5
35510 Butzbach
Germany

Identification of the certified type **An ultrasonic gas meter**
Manufacturers mark: RMG Messtechnik GmbH
Type: USM-GT400

Characteristics See following page(s)

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified type (represented by the samples identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 137-1:2012 "Gas meters"

Accuracy class 0,5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
5 January 2024

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
the Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.





OIML Member State
The Netherlands

OIML Certificate

Number R137/2012-A-NL1-23.04 revision 1
Project number 3711995
Page 2 of 5

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3498274-01 dated 19 May 2023 that includes 51 pages.
- No. NMI-3498274-01 revision 1 dated 22 May 2023 that includes 51 pages.
- No. NMI-3711995-01 dated 5 January 2024 that includes 23 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented.
In Table 2 & 3 the characteristics of the family of instruments are presented.

Table 1 General characteristics

Destined for the measurement of	Gas volume				
Environmental classes	M2 / E2				
Ambient temperature range	-40 ... +70 °C				
Gas temperature range	-40 ... +80 °C				
Designed for	Non-condensing humidity & condensing humidity				
Orientation	Horizontal, vertical up and vertical down (all orientations)				
Power supply voltage	19,2 – 26,4 V DC				
Software identification	Version number:		Checksum:		
	1.523		1AB7		
	1.527		D031		
	1.529		30CF		
Communication output	RS-485 / RS-232				
Transducers	Type	Frequency [kHz]	Pressure range [bar]	Admissible range of diameter [-]	Beam Angle [°]
	TNG 10-CP	120	0 – 150	DN200-DN1000	circa 10
	TNG 10-CHP	120	0 – 300	DN200-DN1000	
	TNG 20-LP	200	0 – 150	DN200-DN600	
	TNG 20-LHP	200	0 – 300	DN200-DN600	
	TNG 20-SP	200	0 – 150	DN80-DN150	
	TNG 20-SHP	200	0 – 300	DN80-DN150	

Table 2 Meter type, path configuration & installation conditions

The meter consists of a cylindrical spool piece with 6 horizontal direct paths, 2 paths on each level evenly distributed. See configuration below that can be combined in a single housing according to the table below.

GT400	Metrological relevant transducer paths	Flow disturbance	Minimum installation conditions up- and downstream ^[1]
Standard	6	Mild & Severe	5D FC ^[1] 5D USM 3D ^[2]
Standard	6	Mild & Severe	5D FC ^[3] 10D USM 3D ^[2]

- [1] xD = Minimum straight inlet & outlet piping length
FC = Flow conditioner: CPA55E
USM = Ultrasonic gas meter
[2] = A temperature sensor can optionally be placed at least 1,5D downstream of the USM outlet.
[3] FC = Flow conditioner: LP35

Table 3 General characteristics of the family of instruments

Standard (6 Path)					
Diameter Nominal	Internal bore ^[2]		Maximum ^[1] V _{max} [m/s]	Minimum ^[1]	
	minimum [mm]	maximum [mm]		V _t [m/s]	V _{min} [m/s]
3" (80 mm)	73,7	77,9	40,01	2,46	0,31
4" (100 mm)	97,2	102,3	39,10	2,31	0,28
6" (150 mm)	146,4	154,1	39,17	2,51	0,31
8" (200 mm)	193,7	202,7	37,81	1,80	0,29
10" (250 mm)	242,8	254,4	37,77	1,43	0,29
12" (300 mm)	288,8	303,2	37,94	1,41	0,28
14" (350 mm)	317,5	333,3	38,08	1,50	0,30
16" (400 mm)	363,5	381	38,29	1,53	0,31
18" (450 mm)	409,6	428,7	36,24	1,51	0,30
20" (500 mm)	455,6	477,8	34,10	1,46	0,29
24" (600 mm)	547,7	574,6	31,45	1,46	0,29

Remarks regarding table 3

[1] If higher values are chosen for Q_{\min} or Q_t and/or lower values for Q_{\max} , it has to be taken into account that:

- If ratio $Q_{\max} / Q_{\min} \geq 50$ then: $Q_{\max} / Q_t \geq 10$
- If ratio $Q_{\max} / Q_{\min} \geq 5$ and < 50 then: $Q_{\max} / Q_t \geq 5$

[2] Depending on the internal bore, the flow rate in m^3/h can be calculated.
The corresponding flow rates can be calculated as follows:

$$Q = v \cdot \frac{1}{4} \cdot \pi \cdot D^2 \cdot 3600$$

Where:

- Q = flow rate [m^3/h]
- v = velocity [m/s]
- D = internal diameter [m]

Installation conditions:

Installation requirements

Any components which could affect the gas flow must be avoided within the prescribed inlet pipe length. The inlet pipe must be designed as a straight pipe section of a nominal diameter $< 3\%$ difference compared to the gas meter.

The installation requirements, based on accuracy class and mild and/or severe flow disturbance are specified in table 2. The used flow conditioner is of type: CPA55E or LP35.

Working pressure

The spool piece and the transducers can be used up to the specified pressure as given in table 1.

The following working pressure ranges are applicable:

- For a working pressure range (p_{\min}) ≥ 10 bar(g) a calibration on one pressure is sufficient. The working pressure range (p_{\min} and p_{\max}) as given on the name plate is allowed to be within $\frac{1}{2} \cdot p_{\text{fix}}$ and $2 \cdot p_{\text{fix}}$.
- For a working pressure range (p_{\min}) < 10 bar(g) a calibration shall be performed on the two pressures at p_{\min} and p_{\max} as given on the name plate.

The ultrasonic gas meter does not make use of an internal pressure or temperature sensor. The correct fixed density (p_{fix}) and fixed pressure setting (p_{fix}) shall be applied before calibration in the USM setting.

For a maximum working pressure (p_{\max}) of the ultrasonic gas meter above 60 bar(a), a maximum calibration pressure ($p_{\max \text{ calibration}}$) of around 60 bar(a) is sufficient since this is the highest possible test pressure at any test location traceable to (inter)national standards.

Thermowell

A thermowell may be mounted at $\geq 1,5D$ and $\leq 5D$ from the outlet of the meter in unidirectional flow measurement.



OIML Member State
The Netherlands

OIML Certificate

Number R137/2012-A-NL1-23.04 revision 1
Project number 3711995
Page 5 of 5

Alternative welded configuration of the gas meter

The central meter body can be welded directly onto the flanges or to inlet and outlet pipes. The welding may not cause more than a 3% diameter step. The meter shall be installed as stated in "Installation requirements". The central meter body, including welded piping or welded flanges, shall be calibrated as a meter package during the examination for putting into use of the gas meter.

Maintenance

A transducer path pair or a single transducer on a path can be exchanged without deterioration of the metrological performance.

Certificate history:

Revision	Date	Description of the modification
0	22 May 2023	Initial
1	5 January 2024	LP35 FC added