



OIML Member State  
P. R. China



OIML Certificate No.  
R46/2012-A-CN2-26.07

## OIML CERTIFICATE

ISSUED UNDER SCHEME A

**OIML Issuing Authority** Name: National Institute of Metrology, China  
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P. R. China  
Person responsible: Mr. Li Wentao

**Applicant** Name: Holley Technology Ltd.  
Address: No.181 Wuchang Avenue Yuhang District, Hangzhou, P.R.CHINA

**Manufacturer** Name: Holley Technology Ltd.  
Address: No.181 Wuchang Avenue Yuhang District, Hangzhou, P. R. China

**Identification of the certified type** Three-phase smart meter DTSD545 Model:H310-DT  
① 3 × 220/380 V 0.25-5(100) A 50 Hz ② 3 × 220/380 V 0.5-10(100) A 50 Hz  
③ 3 × 230/400 V 0.25-5(100) A 50 Hz ④ 3 × 230/400 V 0.5-10(100) A 50 Hz  
⑤ 3 × 240/415 V 0.25-5(100) A 50 Hz ⑥ 3 × 240/415 V 0.5-10(100) A 50 Hz  
(Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R46 – Edition 2012**

For accuracy class: **Class B (Class 1)**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 27 May 2026

**Li Wentao**  
Director  
National Institute of Metrology, China



No.18, Bei San Huan Dong Lu,  
Chaoyang Dist, Beijing,  
100029, P.R.China  
+86 10 64525646  
oimlia@nim.ac.cn  
www.nim.ac.cn

**OIML Member State**  
**P. R. China**

**OIML Certificate No.**  
**R46/2012-A-CN2-26.07**

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R46/2012-CN2-26.07 dated 27 May 2026 that includes 16 pages  
and the associated OIML test report:

No. SIMT2026F14-41-004 dated 20 April 2026 that includes 111 pages

### OIML Certificate History

Revision No.	Date	Description of the modification
0	27 May 2026	Initial issue

### Characteristics of the Instrument

Meter type (electromechanical/static):	static
Accuracy class:	Class B (Class 1)
Nominal voltage, $U_{nom}$ :	3×220/380 V, 3×230/400 V, 3×240/415 V
Nominal frequency, $f_{nom}$ :	50 Hz
Maximum current, $I_{max}$ :	100 A
Transitional current, $I_{tr}$ :	0.5 A, 1 A
Minimum current, $I_{min}$ :	0.25 A, 0.5 A
Starting current, $I_{st}$ :	0.02 A, 0.04 A
Connection mode (phases, wires, elements):	Direct-connected Three phases four wires, three elements
Direction of energy flow /registers:	Two-register, bi-directional
Register multiplier:	1
Meter constant:	1000 imp/kWh
Specified clock frequencies	32.768 kHz
Indoor / Outdoor:	Indoor
IP Rating:	IP54
Terminal arrangement (e.g.: BS, DIN):	DIN
Insulation protection class:	II
Lower specified temperature:	-40 °C
Upper specified temperature:	+70 °C
Humidity class:	H1
Tilt / Mounting position:	Any position is allowed
Hardware version(s):	H310-AA1-01-M03
Software version(s):	Version:HMS3XX25121G5G000096400251121110 Checksum:B02F8E58

Remarks:

Communication interfaces: 1 optical port, 1 RS485 port