



OIML Certificate

OIML Member State

The Netherlands



Number R46/2012-A-NL1-24.02 revision 0 Project number 3706256 Page 1 of 3

Issuing authority NMi Certin B.V.

Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer

Ningbo Sanxing Medical & Electric Co., Ltd.

No.1166 Mingguang North Road

Yinzhou Industrial Zone, Jiangshan Town, Yinzhou District

Ningbo City, Zhejiang Province

China

Identification of the

An Active electrical energy meter

certified type

Type: \$34U18 (6A)

Characteristics See following page(s)

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 46-1/2: 2012 "title OIML Recommendation"

Accuracy class



This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

2 April 2024

Certification Board

NMi Certin B.V. Thijsseweg 11 2629 JA Delft the Netherlands T +31 88 636 2332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.











OIML Member State The Netherlands



Number R46/2012-A-NL1-24.02 revision 0 Project number 3706256 Page 2 of 3

OIML Certificate

The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMi-2534035-01 dated 14 January 2021 that includes 56 pages;
- No. NMi-2534035-02 dated 14 January 2021 that includes 12 pages;
- No. NMi-3706256-01 dated 2 April 2024 that includes 10 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented. The construction of the measuring instrument is recorded in the Documentation folder no. R46-2012-A-NL1-24.02-1.

Table 1 General characteristics

| General characteristics | | |
|--|--|--|
| Meter type Static | | |
| Connection mode (phase, wires, elements) | 3p, 4w, 3e | |
| Direction of energy flow / registers | Two-registers, bi-directional | |
| Terminal arrangement | DIN | |
| Protective class | Category 2 | |
| Environmental application | | |
| Ambient temperature range | -10 °C to +75 °C | |
| Humidity class | H1 | |
| IP Rating / environmental use | IP54 | |
| Meter quantities | | |
| Nominal voltage (U_{nom}) | ominal voltage (<i>U</i> _{nom}) 3x133/230V3x230/400V | |
| Nominal frequency (f _{nom}) 60 Hz | | |
| Maximum current (I _{max}) 6 A | | |
| Transitional current (I_{tr}) 0,075 A ($I_n = 1.5$ A) | | |
| Minimum current (/ _{min}) | 0,015 A | |
| Starting current (I _{st}) | 0,0015 A | |
| Meter constant | 10.000 imp./kWh | |
| Product version | | |
| Hardware version | S34U18 S09.Y2.J1 M20 S34U18 S09.Y2.J2 | |
| Software identification | P3742 2020-08-29 checksum: D766 P7301 2023-12-14 checksum: 3B6C | |





OIML Member State The Netherlands



Number R46/2012-A-NL1-24.02 revision 0 Project number 3706256 Page 3 of 3

OIML Certificate

Certificate history:





| Revision | Date | Description of the modification |
|----------|------------|---|
| Initial | 02-04-2024 | Update of certificate R46-2012-A-NL1-21.04. Testing due to addition of communication circuits and power control relay |
| | | |









