

OIML Member State
The Netherlands

Number R 49/2006-A-NL1-26.02 revision 0
Project number 3719708
Page 1 of 4

Issuing authority NMI Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer Elster Water Metering Ltd.
130 Camford Way, Sundon Park, Luton
Bedfordshire, LU3 3AN
United Kingdom

Identification of the certified type A volumetric **water meter**
Type: V100, V110, V200 and V210 with nominal capacity of 36 rev/liter

Characteristics See following page(s)

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 49-1: 2006 "Water meters intended for the metering of cold potable water and hot water"

Accuracy class 2

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
13 April 2026

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
the Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMI-3719708-01 dated 13 April 2026 that includes 41 pages.

Characteristics of the measuring instrument

In Table 1 and 2 the general characteristics of the measuring instrument are presented. In Table 3 and 4 the characteristics of the family of instruments are presented. The construction of the measuring instrument is recorded in the Documentation folder no. 3719708-1.

Table 1 General characteristics

Measuring principle	Volumetric
Accuracy class	2
Temperature range ambient	-25 °C / +55 °C
Water temperature class	T30 (+0,1 °C / +30 °C)
Maximum admissible pressure (MAP)	1,6 MPa (16 bar)
Orientation	All positions (Horizontal, vertical or diagonal)
Flow profile sensitivity class	U0 and D0 (0 x DN upstream and 0 x DN downstream)
Reverse flow	The sensor is not intended to measure reverse flow
Pressure loss class	Δp 63 (0,63 bar)

Table 2 Additional characteristics for the Hybrid III, Hybrid IV and Hybrid V registers

Environmental class	M1 / O (installed outdoors)		
Electromagnetic environment	E1		
Power supply	Non-replaceable battery (2.5 – 3.75 V)		
Software identification*	Register Name	Version Number	Checksum
	Hybrid III	V03	2A5E
	Hybrid IV	V60.00	3E 77 3A D6
		V60.01	5D D3 A0 2F
		V60.02	32 2A C1 1D
	Hybrid V	V60.03	A4 97 40 30
		V61.00	29 06 EA 26
V61.01	E2 60 10 41		

*) The software version and checksum are displayed on the meter during the circular sequence.

Table 3 General characteristics of the family of instruments

Meter name	Revs/Liter [RPL]	Flow rates [m ³ /h]				Ratio Q3/Q1
		Minimum Q1	Transitional Q2	Permanent Q3	Overload Q4	
V100	36	0.00625	0.010	2.5	3.125	400
V110	36	0.00625	0.010	2.5	3.125	400
V200	36	0.00625	0.010	2.5	3.125	400
V210	36	0.00625	0.010	2.5	3.125	400

Please note that the flow rates Q1, Q2, Q3 and Q4 can be freely chosen as long as:

- Values Q3 and ratio Q3/Q1 are selected from paragraph 3.1 of OIML R49-1: 2006(E);
- Values mentioned for Q1 and Q2 are minimum values and the ratio Q2/Q1 = 1,6;
- Values mentioned for Q3 and Q4 are maximum values and the ratio Q4/Q3 = 1,25;
- The ratio Q3/Q1 is at least 10.

Table 4 General characteristics of the indicating device

Meter name	Register Variant	Indicating range (minimum value) [m ³]	Verification scale interval (minimum resolution) [m ³]
V100	4x4 Register	9999,99999	0,00001
	5x3 Register ¹	99999,9999	0,0001
V110	4x4 Register	9999,99999	0,00001
	5x3 Register ¹	99999,9999	0,0001
V200	Mechanical	99999,99998	0,00002
	Hybrid III, IV or V Register	999999,99999	0,00001
V210	Mechanical	99999,99998	0,00002
	Hybrid III, IV or V Register	999999,99999	0,00001

¹ The 5x3 register shall only be used if the meter has a Q1 greater than 0,013 m³/h.



OIML Member State
The Netherlands

OIML Certificate

Number R 49/2006-A-NL1-26.02 revision 0
Project number 3719708
Page 4 of 4

Production locations

The measuring instrument is produced at one of the following production locations:

- ELSTER WATER METERING LTD
130 Camford Way, Sundon Park, Luton, Bedfordshire, LU3 3AN,
United Kingdom
- ELSTER S.R.O
Dr. Alberta Schweitzera 194, 916 01 Stara Tura
Slovakia
- ELSTER MEDICIÓN S.A.U
Pol. Masti-Loidi, 13 - CP. 20100 Errenteria
Spain
- GEORGE KENT (MALAYSIA) BERHAD
1115, Blok A, Jalan Puchong, Taman Meranti Jaya,
47120 Puchong Selangor Darul Ehsan
Malaysia

Certificate history:

Revision	Date	Description of the modification
0	13 April 2026	Initial issue