



**OIML Member State**  
Czech Republic

**OIML Certificate No.**  
R49/2013-A-CZ1-24.02  
Revision 3

## OIML CERTIFICATE ISSUED UNDER SCHEME A

### OIML Issuing Authority

Name: Czech Metrology Institute  
Address: Okružní 31, 638 00 Brno, Czech Republic

Person responsible: Jan Kalandra

### Applicant

Name: GWF AG  
Address: Obergrundstrasse 119, CH-6005 Luzern, Switzerland

### Manufacturer

Name: GWF AG  
Address: Obergrundstrasse 119, CH-6005 Luzern, Switzerland

### Identification of the certified type *(the detailed characteristics will be defined in the additional pages)*

Water meter – ultrasonic, dry dial  
sonico NANO

### Designation of the module *(if applicable)*

-

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 49

Edition (year): 2013

For accuracy class (if applicable): 2

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

- No. 0511-ER-V124-25 dated 11 March 2026 that includes 39 pages.
  - Test report No. 6015-PT-P0022-25 that includes 74 pages including annex 1
  - Test report No. 6011-PT-SW011-25 that includes 7 pages including annex 1
  - Test report No. 6011-PT-SW019-24 that includes 7 pages including annex 1.
  - Test report No. 6015-PT-P5006-24 that includes 206 pages including annex 1 – 2.
  - Test report No. 6011-PT-SW008-24 that includes 7 pages including annex 1 – 2.
  - Test report No. 6015-PT-P0003-26 that includes 113 pages including annex 1 – 2.
  - Test report No. 6011-PT-SW033-25 that includes 7 pages including annex 1 – 2.

The technical documentation relating to the identified type is contained in documentation file:  
0511-UL-V124-25

#### OIML Certificate History

Revision No.	Date	Description of the modification
-	30 May 2024	Issuing certificate
Revision 1	20 September 2024	Added new SW
Revision 2	12 June 2025	Added new sizes DN25 to DN40 and new SW and CRC checksum versions
Revision 3	11 March 2026	Addition of ratio for water meter DN25 from R630 to R1000, new SW version and addition of module NB-IoT and adding mechanical environment class according to OIML R 49-1:2013: B, O

#### The OIML Issuing Authority

RNDr. Pavel Klenovský  
Head of Certification Body

Date: 11 March 2026



**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

### Measuring system description

The water meters type sonico NANO are designed to measure, memorise and display the volume at metering conditions of water passing through the measurement transducer.

The water meters type sonico NANO are ultrasonic water meters with an electronic indicating device.

The water meters type sonico NANO consist of a cast brass body with connecting screw threads, one pair of ultrasonic transducers and the electronic indicating device. The electronic indicating device is formed by LCD display shown volume and flow. The water meter displays the volume resolution of  $0.00001 \text{ m}^3$  on the digital display. Water meter is without any buttons with LCD display. Legally non-relevant part of communication with meter is possible by NFC sensor connected on the register.

Ultrasonic water meter has a separation of software. Non-legally relevant parts have no inadmissible influence on legally relevant software, measured data or specific parameters.

The version of SWs and CRCs are displayed in the auto-rounding menu on LCD display in the time period in the form:

- CRC of legally relevant part
- SW version of legally relevant part

The water meters type sonico NANO displays the indication of each volume on the display every two minutes – separately delivered volume for reverse flow and separately delivered volume for forward flow. The permanently shown delivered volume is the difference between two delivered volumes (for reverse and forward flow).

The water meters type sonico NANO can be equipped by impulse module which is not part of this certificate.

The water meters type sonico NANO can be assembled by an internal NB-IoT communication module, which was not certified itself, but has no influence on metrological parameters of water meters.

The water meters type sonico NANO are by powered mains battery 3.6V.

The water meters shall be installed to operate in any positions.

### Marking and inscriptions

The water meters types sonico NANO shall be clearly and indelibly marked with the following information:

- Water meter type
- Unit of measurement ( $\text{m}^3$ ) (on display)
- Numerical value  $Q_3$  in  $\text{m}^3/\text{h}$  ( $Q_3 \times \times$ ) and the ratio  $Q_3 / Q_1$ ,
- Manufacturer's name, registered trade name or registered trade mark
- Year of manufacture, two last digits of the year of manufacture, or the month and year of manufacture
- Serial number (as near as possible to the indicating device)
- Direction of flow, by means of an arrow (on display)
- Maximum admissible pressure (MAP  $\times \times$ )
- The temperature class ( $T \times \times$ )
- The pressure loss class ( $\Delta p \times \times$ )
- The installation sensitivity class ( $U \times D \times$ )
- Environmental classification (M)
- Electromagnetic environmental class (E2)
- Type approval sign according to national regulations

These markings shall comply with the requirements of OIML R 49 and shall be visible without dismantling the water meter after the instrument has been placed on the market or put into use.

### Characteristics

Basic technical data of water meters types sonico NANO:

Manufacturer:	GWF AG				
Model name:	sonico Nano				
Nominal diameter:	15	20	25	32	40
Type details:					
$Q_1$ [m <sup>3</sup> /h]:	Flowrates and ratios are shown in Table <i>Basic metrological data</i> (flowrates)				
$Q_2$ [m <sup>3</sup> /h]:					
$Q_3$ [m <sup>3</sup> /h]:					
$Q_4$ [m <sup>3</sup> /h]:					
$Q_3/Q_1$ :					
$Q_2/Q_1$ :	1.6				
$Q_4/Q_3$ :	1.25				
Measuring principle:	ultrasonic				
Accuracy class:	2				
Maximum permissible error for the lower flowrate zone (MPE <sub>l</sub> ):	±5 %				
Maximum permissible error for the upper flowrate zone (MPE <sub>u</sub> ):	±2 %				
Temperature class:	T50				
Water pressure class:	MAP16				
Pressure loss class:	Q <sub>3</sub> 1.6 Δp10 Q <sub>3</sub> 2.5 Δp16	Q <sub>3</sub> 2.5 Δp16 Q <sub>3</sub> 4.0 Δp25	Q <sub>3</sub> 6.3 Δp25 Q <sub>3</sub> 10 Δp40	Q <sub>3</sub> 6.3 Δp16 Q <sub>3</sub> 10 Δp25	Q <sub>3</sub> 10 Δp10 Q <sub>3</sub> 16 Δp16
Reverse flow:	designed to measure				
Mechanical environment class:	B, O, M				
Electromagnetic environment:	E2				
Temperature range ambient:	-25 °C / 70 °C				
Maximum admissible temperature [°C]:	50				
Maximum admissible pressure [MPa]:	1.6				
Orientation limitation:	any				
Indicating range – testing mode/user mode [m <sup>3</sup> ]:	9 999 / 999 999				
Resolution of the indicating device testing mode/user mode [m <sup>3</sup> ]:	0.00001 / 0.001				
Resolution of the device for rapid testing [pulse/dm <sup>3</sup> ]:	100				
Resolution of the indicating device for rapid testing [m <sup>3</sup> ]:	0.000001				
EUT testing requirements (OIML R 49-2:2024, 8.1.8):					
Category:	Ultrasonic water meters				
Case:	B				

Installation details:					
Connection type (screw thread):	NPSM or G type 3/4", 7/8", 1"	NPSM or G type 7/8", 1", 1 1/4"	NPSM or G type 1 1/4", 1 1/2"		NPSM or G type 2"
Minimum straight length of inlet pipe [mm]:	0				
Minimum straight length of outlet pipe [mm]:	0				
Flow profile sensitivity class:	U0D0				
Flow conditioner (details if required):	No				
Mounting:	-				
Orientation:	any				
Other relevant information:	-				
Length [mm]:	≥105	≥105	≥150	≥150	≥150
Reed switch power supply ( $U_{max} / I_{max}$ ):	-				
Reed switch K-factor (impulse / L):	-				
Installation details (electrical):					
Wiring instructions:	-				
Mounting arrangement:	-				
Orientation limitations:	-				
Power supply:					
Type (battery, mains AC, mains DC):	Non replaceable battery				
$U_{max}$ (V):	3.6				
$U_{min}$ (V):	1.9				
Frequency:	-				
Minimum battery life time [years]:	16 years				
Software version (of legally relevant SW):	0.6.28; 1.0.1; 2.7.2; 3.2.2				
CRC checksum (of legally relevant SW):	0x854EBACF, 0x13372073; 0x3A681C19, 0x1d901C40; 0x3F790676; 0x0AB7802F; 0x01F76A0B				
Ancillary devices (not certified):					
Reed sensor power supply ( $U_{max} / I_{max}$ ):	-				
Type	-				
Power supply ( $U_{max} / I_{max}$ ) <sup>2</sup> :	-				
K-factor [pulse/Litres]:	-				
Reed sensor power supply ( $U_{max} / I_{max}$ ):	-				
Other specification of software:					
Specific requirements for embedded software for built-for-purpose measuring instrument (type P)					
Extension I1: Water meters					
Extension S: Software separation					
Extension D: Download of Legally Relevant Software					
Further information specified by the manufacture (not certified)					
-					

<sup>1</sup> The ratio  $Q_3 / Q_1$  shall be chosen according to paragraph 4.1.4 of OIML R 49-1:2013

**Basic metrological data (flowrates)**

**Basic metrological data (flowrates)**

Manufacturer:	GWF AG											
Model name:	sonico NANO											
Nominal diameter:	15											
Type details:												
$Q_1$ [m <sup>3</sup> /h]:	0.040	0.063	0.032	0.050	0.025	0.040	0.020	0.031	0.016	0.025	0.013	0.020
$Q_2$ [m <sup>3</sup> /h]:	0.064	0.100	0.051	0.080	0.041	0.064	0.032	0.050	0.026	0.040	0.021	0.032
$Q_3$ [m <sup>3</sup> /h]:	1.600	2.500	1.600	2.500	1.600	2.500	1.600	2.500	1.600	2.500	1.600	2.500
$Q_4$ [m <sup>3</sup> /h]:	2.000	3.125	2.000	3.125	2.000	3.125	2.000	3.125	2.000	3.125	2.000	3.125
$Q_3/Q_1$ :	40		50		63		80		100		125	

Manufacturer:	GWF AG											
Model name:	sonico NANO											
Nominal diameter:	15											
Type details:												
$Q_1$ [m <sup>3</sup> /h]:	0.010	0.016	0.008	0.013	0.006	0.010	0.005	0.008	0.004	0.006	0.003	0.005
$Q_2$ [m <sup>3</sup> /h]:	0.016	0.025	0.013	0.020	0.010	0.016	0.008	0.013	0.006	0.010	0.005	0.008
$Q_3$ [m <sup>3</sup> /h]:	1.600	2.500	1.600	2.500	1.600	2.500	1.600	2.500	1.600	2.500	1.600	2.500
$Q_4$ [m <sup>3</sup> /h]:	2.000	3.125	2.000	3.125	2.000	3.125	2.000	3.125	2.000	3.125	2.000	3.125
$Q_3/Q_1$ :	160		200		250		315		400		500	

Manufacturer:	GWF AG			
Model name:	sonico NANO			
Nominal diameter:	15			
Type details:				
$Q_1$ [m <sup>3</sup> /h]:	0.0025	0.004	0.003	0.0025
$Q_2$ [m <sup>3</sup> /h]:	0.004	0.006	0.005	0.004
$Q_3$ [m <sup>3</sup> /h]:	1.600	2.500	2.500	2.500
$Q_4$ [m <sup>3</sup> /h]:	2.000	3.125	3.125	3.125
$Q_3/Q_1$ :	630		800	

Manufacturer:	GWF AG											
Model name:	sonico NANO											
Nominal diameter:	20											
Type details:												
$Q_1$ [m <sup>3</sup> /h]:	0.063	0.100	0.050	0.080	0.040	0.064	0.031	0.050	0.025	0.040	0.020	0.032
$Q_2$ [m <sup>3</sup> /h]:	0.100	0.160	0.080	0.128	0.064	0.102	0.050	0.080	0.040	0.064	0.032	0.051
$Q_3$ [m <sup>3</sup> /h]:	2.500	4.00	2.500	4.00	2.500	4.00	2.500	4.00	2.500	4.00	2.500	4.00
$Q_4$ [m <sup>3</sup> /h]:	3.125	5.00	3.125	5.00	3.125	5.00	3.125	5.00	3.125	5.00	3.125	5.00
$Q_3/Q_1$ :	40		50		63		80		100		125	

Manufacturer:	GWF AG											
Model name::	sonico NANO											
Nominal diameter:	20											
Type details:												
$Q_1$ [m <sup>3</sup> /h]:	0.016	0.025	0.013	0.020	0.010	0.016	0.008	0.013	0.006	0.010	0.005	0.008
$Q_2$ [m <sup>3</sup> /h]:	0.025	0.040	0.020	0.032	0.016	0.026	0.013	0.020	0.010	0.016	0.008	0.013
$Q_3$ [m <sup>3</sup> /h]:	2.500	4.00	2.500	4.00	2.500	4.00	2.500	4.00	2.500	4.00	2.500	4.00
$Q_4$ [m <sup>3</sup> /h]:	3.125	5.00	3.125	5.00	3.125	5.00	3.125	5.00	3.125	5.00	3.125	5.00
$Q_3/Q_1$ :	160		200		250		315		400		500	

Manufacturer:	GWF AG			
Model name:	sonico NANO			
Nominal diameter:	20			
Type details:				
$Q_1$ [m <sup>3</sup> /h]:	0.004	0.006	0.005	0.004
$Q_2$ [m <sup>3</sup> /h]:	0.006	0.010	0.008	0.006
$Q_3$ [m <sup>3</sup> /h]:	2.500	4.00	4.00	4.00
$Q_4$ [m <sup>3</sup> /h]:	3.125	5.00	5.00	5.00
$Q_3/Q_1$ :	630		800	1000

Manufacturer:	GWF AG											
Model name::	sonico NANO											
Nominal diameter:	25											
Type details:												
$Q_1$ [m <sup>3</sup> /h]:	0.156	0.250	0.126	0.200	0.100	0.159	0.080	0.125	0.063	0.100	0.050	0.080
$Q_2$ [m <sup>3</sup> /h]:	0.252	0.400	0.202	0.320	0.160	0.254	0.126	0.200	0.101	0.160	0.080	0.128
$Q_3$ [m <sup>3</sup> /h]:	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00
$Q_4$ [m <sup>3</sup> /h]:	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50
$Q_3/Q_1$ :	40		50		63		80		100		125	

Manufacturer:	GWF AG											
Model name::	sonico NANO											
Nominal diameter:	25											
Type details:												
$Q_1$ [m <sup>3</sup> /h]:	0.039	0.063	0.032	0.050	0.025	0.040	0.020	0.032	0.016	0.025	0.013	0.020
$Q_2$ [m <sup>3</sup> /h]:	0.063	0.100	0.050	0.080	0.040	0.064	0.032	0.051	0.025	0.040	0.020	0.032
$Q_3$ [m <sup>3</sup> /h]:	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00
$Q_4$ [m <sup>3</sup> /h]:	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50
$Q_3/Q_1$ :	160		200		250		315		400		500	

Manufacturer:	GWF AG					
Model name:	sonico NANO					
Nominal diameter:	25					
Type details:						
$Q_1$ [m <sup>3</sup> /h]:	0.010	0.016	0.008	0.013	0.006	0.010
$Q_2$ [m <sup>3</sup> /h]:	0.016	0.025	0.013	0.020	0.010	0.016
$Q_3$ [m <sup>3</sup> /h]:	6.30	10.00	6.30	10.00	6.30	10.00
$Q_4$ [m <sup>3</sup> /h]:	7.875	12.50	7.875	12.50	7.875	12.50
$Q_3/Q_1$ :	630		800		1000	


Manufacturer:	GWF AG												
Model name::	sonico NANO												
Nominal diameter:	32												
Type details:													
$Q_1$ [m <sup>3</sup> /h]:	0.156	0.250	0.126	0.200	0.100	0.159	0.080	0.125	0.063	0.100	0.050	0.080	
$Q_2$ [m <sup>3</sup> /h]:	0.252	0.400	0.202	0.320	0.160	0.254	0.126	0.200	0.101	0.160	0.080	0.128	
$Q_3$ [m <sup>3</sup> /h]:	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	
$Q_4$ [m <sup>3</sup> /h]:	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	
$Q_3/Q_1$ :	40		50		63		80		100		125		

Manufacturer:	GWF AG												
Model name::	sonico NANO												
Nominal diameter:	32												
Type details:													
$Q_1$ [m <sup>3</sup> /h]:	0.039	0.063	0.032	0.050	0.025	0.040	0.020	0.032	0.016	0.025	0.013	0.020	
$Q_2$ [m <sup>3</sup> /h]:	0.063	0.100	0.050	0.080	0.040	0.064	0.032	0.051	0.025	0.040	0.020	0.032	
$Q_3$ [m <sup>3</sup> /h]:	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	6.30	10.00	
$Q_4$ [m <sup>3</sup> /h]:	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	7.875	12.50	
$Q_3/Q_1$ :	160		200		250		315		400		500		

Manufacturer:	GWF AG			
Model name:	sonico NANO			
Nominal diameter:	32			
Type details:				
$Q_1$ [m <sup>3</sup> /h]:	0.010	0.016	0.013	0.010
$Q_2$ [m <sup>3</sup> /h]:	0.016	0.025	0.020	0.016
$Q_3$ [m <sup>3</sup> /h]:	6.30	10.00	10.00	10.00
$Q_4$ [m <sup>3</sup> /h]:	7.875	12.50	12.50	12.50
$Q_3/Q_1$ :	630		800	1000

Manufacturer:	GWF AG												
Model name::	sonico NANO												
Nominal diameter:	40												
Type details:													
$Q_1$ [m <sup>3</sup> /h]:	0.250	0.400	0.200	0.320	0.159	0.254	0.125	0.200	0.100	0.160	0.080	0.128	
$Q_2$ [m <sup>3</sup> /h]:	0.400	0.640	0.320	0.512	0.254	0.406	0.200	0.320	0.160	0.256	0.128	0.205	
$Q_3$ [m <sup>3</sup> /h]:	10.00	16.0	10.00	16.0	10.00	16.0	10.00	16.0	10.00	16.0	10.00	16.0	
$Q_4$ [m <sup>3</sup> /h]:	12.50	20.0	12.50	20.0	12.50	20.0	12.50	20.0	12.50	20.0	12.50	20.0	
$Q_3/Q_1$ :	40		50		63		80		100		125		

Manufacturer:	GWF AG												
Model name::	sonico NANO												
Nominal diameter:	40												
Type details:													
$Q_1$ [m <sup>3</sup> /h]:	0.063	0.100	0.050	0.080	0.040	0.064	0.032	0.051	0.025	0.040	0.020	0.032	
$Q_2$ [m <sup>3</sup> /h]:	0.100	0.160	0.080	0.128	0.064	0.102	0.051	0.081	0.040	0.064	0.032	0.051	
$Q_3$ [m <sup>3</sup> /h]:	10.00	16.0	10.00	16.0	10.00	16.0	10.00	16.0	10.00	16.0	10.00	16.0	
$Q_4$ [m <sup>3</sup> /h]:	12.50	20.0	12.50	20.0	12.50	20.0	12.50	20.0	12.50	20.0	12.50	20.0	
$Q_3/Q_1$ :	160		200		250		315		400		500		



Manufacturer:	GWF AG			
Model name:	sonico NANO			
Nominal diameter:	40			
Type details:				
$Q_1$ [m <sup>3</sup> /h]:	0.016	0.024	0.020	0.016
$Q_2$ [m <sup>3</sup> /h]:	0.025	0.041	0.032	0.026
$Q_3$ [m <sup>3</sup> /h]:	10.00	16.0	16.0	16.0
$Q_4$ [m <sup>3</sup> /h]:	12.50	20.0	20.0	20.0
$Q_3/Q_1$ :	630		800	1000

**Securing components and verification marks**

The sonico NANO meters have to be sealed by connecting the plastic seal on the plastic meter cover. The plastic seal is part of the body of the water meter in the form of a plastic frame that holds glass of display on the body of the water meter. The water meter cannot be accessed without damaging seal.

Figure: 1 View on water meter types sonico NANO, display, sealing - example



Figure: 2 Water meter types sonico NANO – description of display

