



OIML Certificate

OIML Member State
The Netherlands

Number R51/2006-A-NL1-21.04 revision 2
Project number 3891265
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Bizerba SE & Co. KG
Wilhelm-Kraut-Straße 65
72336 Balingen
Germany

Identification of the
certified type

An **Automatic catchweighing instrument**
Type : WP...

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 51-1:2006 for accuracy class Y(a) or Y(b)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
2 January 2025

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.





OIML Member State
The Netherlands

OIML Certificate

Number R51/2006-A-NL1-21.04 revision 2
Project number 3891265
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-1901630-01 dated 13 July 2018 that includes 54 pages;
- No. NMI-3547281-01 dated 1 May 2023 that includes 20 pages;
- No. NMI-3741825-01 dated 29 August 2024 that includes 15 pages.

Characteristics of the automatic catchweighing instrument

Destined to be used as	catchweigher, weight labeller, weight/price labeller	
Accuracy class	Y(a), Y(b)	
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$	
Minimum capacity	Min $\geq 20 \text{ e}$ for class Y(a) Min $\geq 10 \text{ e}$ for class Y(b)	
Verification scale interval	$e \geq 1 \text{ g}$	
Weighing ranges	Single interval Multi-interval	
Maximum number of scale intervals (single interval)	$n \leq 3000$ divisions	
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions (per partial weighing range)	
Maximum number of partial weighing ranges	2	
Tare	$T \leq -0,2 \text{ Max}$	
Electromagnetic environment class	E2	
Climatic environment	temperature range	$-10 \text{ }^{\circ}\text{C} / +40 \text{ }^{\circ}\text{C}$
	humidity	non-condensing
	intended location	closed
Power supply voltage	400 V AC 50 Hz, or 208-240 V AC 50/60 Hz, or 230 V AC 50/60 Hz	



OIML Certificate

OIML Member State
The Netherlands

Number R51/2006-A-NL1-21.04 revision 2
Project number 3891265
Page 3 of 3

Software identification for program modules ¹⁾	RetailPowerScale	Version	Software ID (Windows)	Software ID (Linux)
		012	6914	7130
		014	0788	2354
	Scale OEM Module	011	3362	1418
		013	7640	0369
Software identification for AD board ²⁾		ADW ID	Software version	Software ID
		047195xxxx ³⁾	004	7195
		106153xxxx ³⁾	010	6153
		115232xxxx ³⁾	011	5232
		301146xxxx ³⁾	030	1446
		144751xxxx ³⁾	014	4751

- ¹⁾ The identification of the non-legally relevant software version can optionally be added to the identification of the legally relevant software version, separated by "::" (example: "013::x.y").
- ²⁾ Alternatively, the software version number and the software-ID can be displayed separately. In this case, the identification of the non-legally relevant software version can optionally be added to the identification of the legally relevant software version separated by "::" (example: "6153 010::xx.xx").
- ³⁾ "x" represents the non-legally relevant part of the software and can be any number between 0...9.

Software:

- The identification number will be shown after tapping the part of display with inscriptions Max, Min, e and the accuracy class at least for 3 seconds.

Revision History

This revision replaces the previous versions.

Revision	Date	Change(s)
0	2021-04-02	Initial issue.
1	2023-05-04	Additional type evaluation report and new software version included.
2	2025-01-02	Addition of an alternative A/D board and a new software version.