

## OIML Member State Switzerland

# OIML Certificate No. R060/2000-A-CH1-25.01

## OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

Name Federal Institute of Metrology METAS

Conformity Evaluation Body METAS-Cert

Address Lindenweg 50, 3003 Bern-Wabern, Switzerland

Person responsible Gulian Couvreur, Head of METAS-Cert

**Applicant** 

Name DIGI SENS Switzerland AG

Address Freiburgstrasse 65, 3280 Murten, Switzerland

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified type

Type SLC NZ

Characteristics  $E_{max}$  1500 kg with  $v_{min}$  0.6 kg

n<sub>LC</sub> 1500 Y 2500 p<sub>LC</sub> 0.8

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R60, edition 2000

for accuracy class C 1.5

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

This document is only valid and reviewable in its electronic form. Please observe the information given on www.metas.ch/ecert

#### OIML-CS Certificate No R060/2000-A-CH1-25.01

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No. Date Including pages

6030-02916 2025-02-11 8

The technical documentation relating to the identified type is contained in documentation file:

Name Date Including pages

R060/2000-A-CH1-25.01\_LERD\_DIGI-SENS 2025-02-11 4

### OIML Certificate History:

Revision No.	Date	Description of the modification
00	2025-02-11	First issue

#### The OIML Issuing Authority CH1

3003 Bern-Wabern, 2025-02-11

Approved by Gulian Couvreur, Head of sector

METAS-Cert



Important note: Apart from the mention of the Certificate's reference number and the

name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type

evaluation report(s) is not permitted, although either may be

reproduced in full