



OIML Certificate

OIML Member State
The Netherlands

Number R60/2021-A-NL1-26.09 revision 1
Project number 3940755
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant

Mettler-Toledo GmbH
Im Langacher 44
8606 Greifensee
Switzerland

Manufacturer

Mettler-Toledo (Changzhou) Precision Instruments Ltd.
No.22, Zhengqiang Road, Xinbei District
Changzhou, Jiangsu
P.R. China

Identification of the
certified type

A **single point load cell**, with strain gauges, equipped with electronics.
Registered trade name : METTLER TOLEDO
Type : SLP86xD (x=1,2,3)

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60-1:2021 for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
29 June 2026

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
the Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R60/2021-A-NL1-26.09 revision 1
Project number 3940755
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3940755-01 dated 15 June 2026 that includes 71 pages;
- No. NMI-3940755-02 dated 15 June 2026 that includes 46 pages;
- No. NMI-3940755-03 dated 15 June 2026 that includes 46 pages;
- No. NMI-3940755-04 revision 1 dated 29 June 2026 that includes 25 pages.

Characteristics of the load cell:

Characterization of load cell capabilities	Digital load cell with data processing
Maximum capacity ($E_{\max}$)	10 kg up to and including 750 kg
Minimum dead load	0 kg
Accuracy Class	C
Maximum number of load cell intervals (n) ⁽¹⁾	10000
Ratio of minimum LC Verification interval ⁽¹⁾ $Y = E_{\max} / v_{\min}$	40000
Ratio of minimum dead load output return ⁽¹⁾ $Z = E_{\max} / (2 * DR)$	10000
Temperature range	-10 °C / +40 °C
Fraction p_{LC}	0,8
Humidity Class	CH
Safe overload	150 % of $E_{\max}$
Recommended excitation	12-24 V DC
Excitation maximum	30 V DC
Transducer material	Stainless steel
Atmospheric protection	Hermetically welded
Electromagnetic environment class	E2
Number of counts for $E_{\max}$	$\geq Y * 5 / p_{LC}$
Software identification	Described in the certificate TC8039

Remark:

1. The characteristics for $n_{\max}$, Y and Z can be reduced separately.

Each load cell produced is provided with an accompanying document with information about its characteristics.



OIML Certificate

OIML Member State
The Netherlands

Number R60/2021-A-NL1-26.09 revision 1
Project number 3940755
Page 3 of 3

The above identified Type (represented by the sample(s) identified in the OIML Test Report) have been found to comply with the additional national requirements established by the United States of America (NIST Handbook 44 and NCWM Publication 14), included in the Utilizer Declaration:

- R 60 OIML-CS rev.2 Additional requirements from the United States Accuracy class III L;
- R 60 OIML-CS rev.2 Additional requirements from the United States Marking requirements.

Revision History

Revision	Date	Change(s)
0	2026-06-15	Initial issue.
1	2026-06-29	Test report NMI-3940755-04 replaced due to correction in it.