OIML Certificate

OIML Member State
The Netherlands

Number R60/2021-A-NL1-26.14 revision 0
Project number 4129610

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Schmidt
Applicant and Manufacturer	Minebea Intec GmbH Meiendorfer Strasse 205 A D-22145 Hamburg Germany
Identification of the certified type	A **compression load cell**, with strain gauges. Registered trade name : Minebea Intec GmbH Type : PR 6261
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60-1:2021 for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
2 October 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
the Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.

OIML Certificate

OIML Member State
The Netherlands

Number R60/2021-A-NL1-26.14 revision 0
Project number 4129610

The conformity was established by the results of tests and examinations provided in the associated reports:
- No. NMi-1900903-01 dated 28 April 2017 that includes 51 pages;
- No. NMi-1900903-02 dated 28 April 2017 that includes 46 pages;
- No. NMi-4129610-01 dated 2 October 2026 that includes 70 pages.

Characteristics of the load cell:

Characterization of load cell capabilities	Analog-passive load cell
Maximum capacity (E_{max})	125 kg up to and including 8 t
Minimum dead load	0 kg
Accuracy Class	C
Rated Output	2,0 mV/V
Maximum number of load cell intervals (n) (1)	3000
Ratio of minimum LC Verification interval (1) $Y = E_{max} / v_{min}$	14000
Ratio of minimum dead load output return (1) $Z = E_{max} / (2 * DR)$	3000
Input impedance	1080 Ω ± 10 Ω
Temperature range	-10 °C / + 40 °C
Fraction p_{LC}	0,7
Humidity Class	CH
Safe overload	150 % of E_{max}
Output impedance	1010 Ω ± 1 Ω
Recommended excitation	10 V AC / DC
Excitation maximum	24 V AC / DC
Transducer material	Stainless steel
Atmospheric protection	Hermetically welded

Remark:
1. The characteristics for n_{max}, Y and Z can be reduced separately.

Each load cell produced is provided with an accompanying document with information about its characteristics.

The above identified Type (represented by the sample(s) identified in the OIML Test Report) have been found to comply with the additional national requirements established by the United States of America (NIST Handbook 44 and NCWM Publication 14), included in the Utilizer Declaration:
- R 60 OIML-CS rev.2 Additional requirements from the United States Accuracy class III L;
- R 60 OIML-CS rev.2 Additional requirements from the United States Marking requirements.

OIML Certificate

OIML Member State
The Netherlands

Number R60/2021-A-NL1-26.14 revision 0
Project number 4129610

Revision History

Revision	Date	Change(s)
0	2026-10-02	Initial issue.