OIML Certificate

OIML Member State
The Netherlands

Number R60/2021-A-NL1-26.19 revision 0
Project number 4019966

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Schmidt
Applicant and Manufacturer	3S Fabrications (Pvt) Ltd. (Group Four Transducers Inc.) Etiyawala, Venus Mills, Daluwela, Dankotuwa 61130 Sri Lanka
Identification of the certified type	A **tension load cell**, with strain gauges. Registered trade name : 3S Fabrications (Pvt) Ltd. (Group Four Transducers Inc.) Type : HRS
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60-1:2021 for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
13 October 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.

OIML Certificate

OIML Member State
The Netherlands

Number R60/2021-A-NL1-26.19 revision 0
Project number 4019966

The conformity was established by the results of tests and examinations provided in the associated report(s):
- No. NMi-4019966-01 dated 13 October 2026 that includes 50 pages.

Characteristics of the load cell:

The construction of the load cell is recorded in the Documentation folder no. TC13324-1.

Characterization of load cell capabilities	Analog-passive load cell
Maximum capacity (E_{max})	250 kg up to and including 1250 kg
Minimum dead load	0 kg
Accuracy Class	C
Rated Output	2,8 mV/V
Maximum number of load cell intervals (n) (1)	3000
Ratio of minimum LC Verification interval (1) $Y = E_{max} / v_{min}$	10000
Ratio of minimum dead load output return (1) $Z = E_{max} / (2 * DR)$	3000
Input impedance	386 Ω ± 30 Ω
Temperature range	-10 °C / + 40 °C
Fraction p_{LC}	0,7
Humidity Class	NH
Safe overload	200 % of E_{max}
Output impedance	350 Ω ± 5 Ω
Recommended excitation	5 V AC / DC
Excitation maximum	15 V AC / DC
Transducer material	Stainless steel
Atmospheric protection	Silicone sealing

Remark:
1. The characteristics for n_{max}, Y and Z can be reduced separately.

Each load cell produced is provided with an accompanying document with information about its characteristics.

OIML Certificate

OIML Member State
The Netherlands

Number R60/2021-A-NL1-26.19 revision 0
Project number 4019966

The above identified Type (represented by the sample(s) identified in the OIML Test Report) have been found to comply with the additional national requirements established by the United States of America (NIST Handbook 44 and NCWM Publication 14), included in the Utilizer Declaration:

- R 60 OIML-CS rev.2 Additional requirements from the United States Accuracy class III L;
- R 60 OIML-CS rev.2 Additional requirements from the United States Marking requirements.

Revision History

Revision	Date	Change(s)
0	2026-10-13	Initial issue.