



# OIML Certificate

**OIML Member State**  
The Netherlands

Number R61/2017-A-NL1-26.02 revision 0  
Project number 3891402  
Page 1 of 3

Issuing authority NMI Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer Accrapak Systems Limited  
Unit 57 Burtonwood Industrial Centre  
Warrington, WA5 4HX  
United Kingdom

Identification of the certified type An **Automatic gravimetric filling instrument**  
Type : 600DC

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 61-1:2017** for reference class Ref 0,2

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
11 March 2026

Certification Board

NMI Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R61/2017-A-NL1-26.02 revision 0  
Project number 3891402  
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated report:

- No. NMI-3891402-01 dated 11 March 2026 that includes 31 pages.

## Characteristics of the automatic gravimetric filling instrument

The indicator:

Producer	Type	Certificate number
SysTec Systemtechnik und Industrieautomation GmbH	IT1 / IT3 series	R76-2006-NL1-16.57

Load cell:

Producer	Type	Certificate number
TEDEA-HUNTLEIGH	1260	R60/2000-NL-02.34

Rated minimum fill with automatic tare-setting or automatic zero-setting as part of every weighing cycle:

Average number of loads per fill:	1			
Accuracy class:	X(0,2)	X(0,5)	X(1)	X(2)
d [g]	Minifill [g]			
0,5	28,0	11,0	5,5	3,0
1	111	22	11	6
2	334	44	22	12
5	1665	335	110	30
10	3330	1330	330	110
20	6660	2660	1340	340
50	25000	6650	3350	1650
100	50000	20000	6700	3300
200	100000	40000	20000	6600
≥500	500 d	200 d	100 d	50 d



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R61/2017-A-NL1-26.02 revision 0  
Project number 3891402  
Page 3 of 3

## Revision History

Revision	Date	Change(s)
0	2026-03-11	Initial issue.