



Member State
Switzerland

OIML Certificate No.
R076/2006-A-CH1-18.09 Rev. 3

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

Name

Federal Institute of Metrology METAS
Conformity Evaluation Body METAS-Cert

Address

Lindenweg 50, 3003 Bern-Wabern, Switzerland

Person responsible

Gulian Couvreur, Head of METAS-Cert

Applicant

Name

Mettler-Toledo GmbH

Address

Im Langacher 44, 8606 Greifensee, Switzerland

Manufacturer

The manufacturer of the certified pattern is the Applicant

Identification of the certified type

Type

XPR Micro Balances

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R 76-1, edition 2006

for accuracy class

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Metrological characteristics

Accuracy class	Ⓢ
Max	$\leq 10.1 \text{ g}$
e	$= 1 \text{ mg}$
d	$\geq 0.1 \text{ }\mu\text{g}$
n	$\leq 10 \text{ } 100$
Tare-weighing range	Until 100 % of Max
Temperature range	10 °C / 30 °C

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
CH-MIER-26001-02	2026-09-03	10

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
CH-W1-16022-04_LERD	2026-09-03	4

OIML Certificate History:

Revision No.	Date	Description of the modification
0	2018-10-01	First issue, replaces OIML MAA Certificate No R76/2006-CH1-16.02
1	2026-05-12	New Software Version und Software Evaluation.
2	2026-05-19	Revisions of Type Evaluation Report and LERD
3	2026-09-03	New Software Version

The OIML Issuing Authority CH1

3003 Bern-Wabern, 2026-09-03

Approved by

Gulian Couvreur, Head of sector
METAS-Cert



Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full