



Member State
Switzerland

OIML Certificate No.
R076/2006-A-CH1-24.01 Rev. 1

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

<i>Name</i>	Federal Institute of Metrology METAS Conformity Evaluation Body METAS-Cert
<i>Address</i>	Lindenweg 50, 3003 Bern-Wabern, Switzerland
<i>Person responsible</i>	Gulian Couvreur, Head of METAS-Cert

Applicant

<i>Name</i>	Ishida Co., Ltd.
<i>Address</i>	44 Sanno-Cho, Shogoin, Sakyo-Ku, Kyoto, 606-8392, Japan

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified type

AI Family

Type **WM-AI, IP-AI, WM-MICRO**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R 76-1, edition 2006

for accuracy class



This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Metrological characteristics

Accuracy class

III

Model

WM-AI

<i>Max</i>	3 / 6 kg	6 / 15 kg
<i>Min</i>	20 g	40 g
<i>e</i>	1 / 2 g	2 / 5 g
<i>n</i>	3 000	3 000
<i>Temperature range</i>	5 °C / 35 °C	5°C / 35 °C
<i>T</i>	- 2.999 kg	- 5.998 kg

Model

IP-AI

<i>Max</i>	3 / 6 kg	6 / 15 kg
<i>Min</i>	20 g	40 g
<i>e</i>	1 / 2 g	2 / 5 g
<i>n</i>	3 000	3 000
<i>Temperature range</i>	0°C / 40°C	0°C / 40°C
<i>T</i>	- 2.999 kg	- 5.998 kg

Model

IP-AI with weighting platform S

<i>Max</i>	30 kg	60 kg	120 kg
<i>Min</i>	100 g	200 g	400 g
<i>e</i>	> 5 g	> 10 g	> 20 g
<i>n</i>	≤ 6 000	≤ 6 000	≤ 6 000
<i>Temperature range</i>	0 °C / 35°C	0 °C / 35°C	0°C / 35°C
<i>T</i>	- 29.995 kg	- 59.99 kg	- 119.98 kg

Model

MICRO

<i>Max</i>	3 / 6 kg	6 / 15 kg
<i>Min</i>	20 g	40 g
<i>e</i>	1 / 2 g	2 / 5 g
<i>n</i>	3 000	3 000
<i>Temperature range</i>	0°C / 40°C or 5°C / 35°C	0°C / 40°C or 5°C / 35°C
<i>T</i>	- 2.999 kg	- 5.998 kg

Annex to the OIML Type Evaluation Report No R076/2006-A-CH1-24.01 Revision 1

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
6030-02871	2025-09-02	16

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
R076/2006-A-CH1-24.01-01_LERD	2025-09-02	6

OIML Certificate History:

Revision No.	Date	Description of the modification
00	2024-07-17	First issue
01	2025-09-02	Correction of n value for S model

The OIML Issuing Authority CH1

3003 Bern-Wabern, 2025-09-02

Approved by

Gulian Couvreur, Head of sector
METAS-Cert



Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full