



Member State
Switzerland

OIML Certificate No.
R076/2006-A-CH1-26.01

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

<i>Name</i>	Federal Institute of Metrology METAS Conformity Evaluation Body METAS-Cert
<i>Address</i>	Lindenweg 50, 3003 Bern-Wabern, Switzerland
<i>Person responsible</i>	Gulian Couvreur, Head of METAS-Cert

Applicant

<i>Name</i>	Mettler-Toledo Instruments (Shanghai) Co. Ltd
<i>Address</i>	No. 589 Gui ping Road, 200233 Shanghai, People's Republic of China

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified type

<i>Type</i>	XPR Micro Balances
-------------	---------------------------

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R 76-1, edition 2006

for accuracy class

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

OIML-CS Certificate No R076/2006-A-CH1-26.01

Metrological characteristics

Accuracy class	I
Max	$\geq 1 \text{ g}, \leq 10.1 \text{ g}$
e	$= 1 \text{ mg}$
d	$\geq 0.1 \mu\text{g}$
n	$\leq 10 \text{ 100}$
Tare-weighing range	Until 100 % of Max
Temperature range	10 °C / 30 °C

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	Date	Including pages
CH-MIER-26003-00	2026-06-15	10

The technical documentation relating to the identified type is contained in documentation file:

Name	Date	Including pages
R076/2006-A-CH1-26.01_LERD	2026-06-15	8

OIML Certificate History:

Revision No.	Date	Description of the modification
00	2026-06-15	First issue

The OIML Issuing Authority CH1

3003 Bern-Wabern, 2026-06-15

Approved by

Gulian Couvreur, Head of sector
METAS-Cert



Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full