



OIML Member State  
P. R. China



OIML Certificate No.  
R76/2006-A-CN2-25.08

## OIML CERTIFICATE

ISSUED UNDER SCHEME A

**OIML Issuing Authority** Name: National Institute of Metrology, China  
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China  
Person responsible: Mr. Li Wentao


**Applicant** Name: Guangdong Province Joytour Technology Co., Ltd.  
Address: Room 601, Building 12, CIMC High Tech Intelligent Manufacturing Center,  
No.3 Chanxing 1st Road, Chencun Town, Shunde District, Foshan City, Guangdong  
Province, P.R.China

**Manufacturer** Name: Guangdong Province Joytour Technology Co., Ltd.  
Address: Room 601, Building 12, CIMC High Tech Intelligent Manufacturing Center,  
No.3 Chanxing 1st Road, Chencun Town, Shunde District, Foshan City, Guangdong  
Province, P.R.China

**Identification of the certified type** Checkout Scale T series (T15,T30) (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76 – Edition 2006**

For accuracy class: 

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 29 December 2025



No.18, Bei San Huan Dong Lu,  
Chaoyang Dist, Beijing,  
100029, P.R.China  
+86 10 64525646  
oiml1a@nim.ac.cn  
www.nim.ac.cn

**Li Wentao**  
Director  
National Institute of Metrology, China

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R76/2006-CN2-25.08 dated 29 December 2025 that includes 25 pages  
and the associated OIML test report:

No. LSmm2025-04789 dated 25 December 2025 that includes 44 pages

### OIML Certificate History

Revision No.	Date	Description of the modification
0	29 December 2025	Initial issue

### Characteristics of the Instrument

Variants of type	T series	
	T15	T30
Accuracy class	III	
Maximum capacity, Max	6 kg/15 kg	15 kg/30 kg
Minimum capacity, Min	40 g	100 g
Verification scale interval, $e$	2 g/5 g	5 g/10 g
Actual scale interval, $d$	2 g/5 g	5 g/10 g
Number of verification scale intervals, $n$	3000/3000	
Weighing range(s)	Multi-interval	
Tare-weighing range	-5.998 kg	-14.995 kg
Temperature range	0 °C ~ 40 °C	
Power supply	AC (100-240) V, 50/60 Hz	
Software identification	V1.00	