



**OIML Member State** 

Germany

# OIML Certificate No. R76/2006-A-DE1-24.03

# OIML CERTIFICATE ISSUED UNDER SCHEME A

# **OIML Issuing Authority**

Name:	Physikalisch-Technische Bundesanstalt,	
	Conformity Assessment Body	
Address:	Bundesallee 100, 38116 Braunschweig, GERMANY	
Person responsible:	DrIng. Prof. h. c. Frank Härtig	

# Applicant

Name:	Mettler-Toledo GmbH
Address:	Im Langacher 44
	CH 8606 Greifensee
	Switzerland

# Manufacturer

Name:	Mettler-Toledo GmbH
Address:	Im Langacher 44
	CH 8606 Greifensee
	Switzerland

**Identification of the certified type** (the detailed characteristics will be defined in the additional pages)

Non-automatic electromechanical weighing instrument Type: IND400 with PBK98x or PFK98x

#### **Designation of the modules** (*if applicable*)

#### Terminal

Type: IND400

The terminal has undergone tests as per the following test reports:

- 1. NMi-3699033-1
- 2. NMi-3699033-2
- 3. NMi-3699033-3

### Weighing module

Type: PBK98x / PFK98x

The weighing module has undergone tests as per the following test reports:

- 1. 20141101.A04.01
- 2. 20141102.A04.01
- 3. 20141103.A04.01
- 4. 20140728.AMNufer

#### Digital load cell

Type: MPGI/MPXI

The digital load cell has undergone tests as per the following test reports:

- 1. 20141088.A04.01
- 2. 20141088.A04.02
- 3. 20141088.A04.03
- 4. 20141088.A04.11
- 5. 20161072.A01.01
- 6. 20161072.A01.02
- 7. 20161072.A01.03
- 8. 1.12-4075006/1

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76

Edition (year): 2006

For accuracy class (if applicable): II, III

# OIML Certificate No. R76/2006-A-DE1-24.03

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. PTB-1.12-4118583 dated 10.04.2025 that includes 15 pages

OIML Certificate	•				
<b>Revision</b> N	No.	Date	Description of the modifica	tion	
0	10.04.	2025	Initial Issuing		
The Issuing Auth Digital signature	ority				
Digital signature					
DrIng. Oliver M	ack				
Member of Confe	ormity Assessment	Body			
Date: 10.04.2025					
Important note:	Apart from the mention of the Certificate's reference number and the name of the				
	OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted,				
	although either ma			mue	
		y oe reproduced h			
			OIML Certifica	ate N	
			R76/2006-A-DE1		

# Identification of the certified type (continued)

Metrological characteristics of the pattern:

Accuracy class		II	III
Maximum capacity Max	-	$\leq$ 3 kg 3 t	$\leq 3 \text{ kg} \dots 60 \text{ t}^{\text{a})}$
Verification scale interval e	g	$\geq 0,1$	
Minimum load Min	-	$\geq$ 50 e <sup>b)</sup>	≥20 e
Number of verification scale intervals	n	≤ 30.000	≤ 10.000
Number of verification scale intervals <sup>c)</sup>	ni	-	≤ 10.000
$Max / e_1^{d}$	-	-	≤ 30.000
Tare balancing range (subtractive)	-	$\leq 100\% \bullet Max$	
Initial zero-setting range	-	$\leq$ 20% • Max	
Temperature range	-	$0^{\circ}C / +40^{\circ}C^{e}$	-10°C / +40°C <sup>f)</sup>

<sup>b)</sup> if e=10d; e can be replaced by d=>  $\geq$  5 e

<sup>c)</sup> For multiple range instruments and multi-interval instruments

<sup>d)</sup> Only for multi-interval instruments

 <sup>e)</sup> Accuracy class II Accuracy class III (PBK987/PBK989 [A3/A6])

<sup>f)</sup> Accuracy class III