




| | | |
|--|---|---|
|  |  DANAK PROD Reg.no. 7026 | FORCE  Certification |
| OIML Member State Denmark | OIML Certificate No. R76/2006-A-DK2-26.03 | |
| OIML CERTIFICATE ISSUED UNDER SCHEME A | | |
| OIML Issuing Authority Name: FORCE Certification A/S Address: Park Allé 345, 2605 Brøndby, Denmark Person responsible: Per Rafn Crety | | |
| Applicant Name: Tscale Electronics Mfg. (Kunshan) Co., Ltd. Address: No. 99 Jingwei Road, Zhoushi, Kunshan, Jiangsu China | | |
| Manufacturer Tscale Electronics Mfg. (kunshan) Co., Ltd. | | |
| Identification of the certified type <i>(the detailed characteristics will be defined in the additional pages)</i> XD... / XP... | | |
| Designation of the module <i>(if applicable)</i> Non-automatic weighing instrument | | |
| <p>This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):</p> <p>OIML R 76-1, Edition (year): 2006</p> <p>For accuracy class (if applicable): III</p> | | |

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type Examination Report: No. 125-22481.10 dated 10th of April 2025, that includes 102 pages.

Type Evaluation Report: No. 126-21693.20, dated 12th of February 2026, that includes 22 pages.

The technical documentation relating to the identified type is contained in documentation file:

125-22481

126-21693

OIML Certificate History

| Revision No. | Date | Description of the modification |
|---------------------|-----------------------------------|--|
| Initial version | 13 th of February 2026 | - |
| | | |
| | | |
| | | |

Identification, signature and stamp

The OIML Issuing Authority

FORCE Certification A/S

Date: 13th of February 2026

Rasmus Møller Jensen

Certification Manager

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Descriptive annex

Characteristics

Type.....: XP... / XD...
 Accuracy class.....: III
 Weighing range: Single interval, dual-range or dual-interval
 Maximum number of
 Verification Scale Intervals: ≤ 3000
 Maximum capacity (Max).....: 3 kg to 30 kg
 Minimum capacity (Min): $20 \times e_1$
 Verification Scale Interval (e).....: ≥ 0.5 g
 Maximum tare effect: -Max
 Operational temperature: -10°C to $+40^{\circ}\text{C}$
 Mains power supply: 12 VDC supplied from an external mains adapter
 intended for 100-240 VAC, 50/60 Hz; or
 7.4V lithium battery.

Models

| Model | Class | Max | e | d | n |
|---------|-------|------------|-------------|-------------|--------|
| XD-3M | III | 3 kg | 1 g | 1 g | 3000 |
| XD-6M | III | 6 kg | 2 g | 2 g | 3000 |
| XD-15M | III | 15 kg | 5 g | 5 g | 3000 |
| XD-30M | III | 30 kg | 10 g | 10 g | 3000 |
| XD-3MR | III | 1.5 / 3 kg | 0.5 g / 1 g | 0.5 g / 1 g | 2x3000 |
| XD-6MR | III | 3 / 6 kg | 1 g / 2 g | 1 g / 2 g | 2x3000 |
| XD-15MR | III | 6 / 15 kg | 2 g / 5 g | 2 g / 5 g | 2x3000 |
| XD-30MR | III | 15 / 30 kg | 5 g / 10 g | 5 g / 10 g | 2x3000 |
| XP-3M | III | 3 kg | 1 g | 1 g | 3000 |
| XP-6M | III | 6 kg | 2 g | 2 g | 3000 |
| XP-15M | III | 15 kg | 5 g | 5 g | 3000 |
| XP-30M | III | 30 kg | 10 g | 10 g | 3000 |
| XP-3MR | III | 1.5 / 3 kg | 0.5 g / 1 g | 0.5 g / 1 g | 2x3000 |
| XP-6MR | III | 3 / 6 kg | 1 g / 2 g | 1 g / 2 g | 2x3000 |
| XP-15MR | III | 6 / 15 kg | 2 g / 5 g | 2 g / 5 g | 2x3000 |
| XP-30MR | III | 15 / 30 kg | 5 g / 10 g | 5 g / 10 g | 2x3000 |

Software

Pressing “TARE” during self-checking will show the legal software version in the display.

Legal software version is: 2.00.

Non-legal application software version is: v1.xxx(y) where x can be 0-9 and y can be a to z.

The scales have software separation.

Devices

- Zero-setting range: ± 2 % of Max.
- Automatic zero-tracking range: ± 2 % of Max.
- Initial zero-setting range: ± 10 % of Max.
- Semi-automatic subtractive tare
- Price Look Up (XP... only)
- Unit weight look up (XD... only)
- Check Weighing device (XD... only)
- Counting device (XD... only)
- Percentage weighing device (XD... only)
- Totalization device
- Gravity compensation
- Printing device (can be connected)

Interfaces

- RS 232
- Ethernet (Optional)
- RJ11 (optional)
- USB (optional)
- Wi-Fi (optional)
- Bluetooth (optional)