
	 DANAK PROD Reg.no. 7026	FORCE  Certification
OIML Member State Denmark	OIML Certificate No. R76/2006-A-DK2-26.04	
OIML CERTIFICATE ISSUED UNDER SCHEME A		
OIML Issuing Authority Name: FORCE Certification A/S Address: Park Allé 345, 2605 Brøndby, Denmark Person responsible: Per Rafn Crety		
Applicant Name: BALANÇAS MARQUES DE JOSÉ PIMENTA MARQUES, LDA. Address: PARQUE INDUSTRIAL DE CELEIRÓS, LOTE D3 2ª FASE 4705-471 CELEIRÓS BRG BRAGA Portugal		
Manufacturer BALANÇAS MARQUES DE JOSÉ PIMENTA MARQUES, LDA.		
Identification of the certified type <i>(the detailed characteristics will be defined in the additional pages)</i> BM ONE		
Designation of the module <i>(if applicable)</i> Non-automatic weighing instrument		
<p>This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):</p> <p>OIML R 76-1, Edition (year): 2006</p> <p>For accuracy class (if applicable): III</p>		

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type Examination Report: No. 125-28314.10 dated 17th of February 2026, that includes 68 pages.

Type Evaluation Report: No. 125-28314.20, dated 26th of February 2026, that includes 20 pages.

The technical documentation relating to the identified type is contained in documentation file:
125-28314

OIML Certificate History

Revision No.	Date	Description of the modification
Initial version	26 th of February 2026	

Identification, signature and stamp

The OIML Issuing Authority

FORCE Certification A/S

Date: 26th of February 2026

Rasmus Møller Jensen

Certification Manager

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Descriptive annex

Characteristics

Type.....	: BM ONE
Accuracy class.....	: III
Weighing range	: Single interval, multi-interval (dual) or multi-range (dual)
Maximum number of Verification Scale Intervals	: ≤ 3000 or 2×3000
Maximum capacity (Max).....	: 6 kg to 30 kg
Minimum capacity (Min)	: $20 \times e$
Verification Scale Interval (e)	: ≥ 1 g
Maximum tare effect	: $\leq -\text{Max}$
Operational temperature	: 0°C to $+40^\circ \text{C}$
Mains power supply	: 230 VAC, 50/60 Hz

Software

The instrument has software separation and is divided into a legal and non-legal part. The software nomenclature is v.xX.yy. xX represent the legal part of the software while yy represents minor changes of the software. The version of the software is shown during initialisation of the scale. The approved software version is: 0.yy

Devices

- Zero-setting range: ± 2 % of Max.
- Automatic zero-tracking range: ± 2 % of Max.
- Initial zero-setting range: ± 10 % of Max.
- Semi-automatic subtractive tare
- Printing device (integrated)

Interfaces

- Cash Drawer interface (RJ-11 connector)
- USB (for setup purposes)