

		
OIML Member State Denmark		OIML Certificate No. R76/2006-A-DK2-26.07
OIML CERTIFICATE ISSUED UNDER SCHEME A		
OIML Issuing Authority Name: FORCE Certification A/S Address: Park Allé 345, 2605 Brøndby, Denmark Person responsible: Per Rafn Crety		
Applicant Name: Avery Berkel Address: Foundry Lane, Smethwick, West Midlands B66 2LP UNITED KINGDOM		
Manufacturer Avery Berkel		
Identification of the certified type <i>(the detailed characteristics will be defined in the additional pages)</i> X One Series		
Designation of the module <i>(if applicable)</i> Non-automatic price-computing weighing instrument		
This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML): OIML R 76-1, Edition (year): 2006 For accuracy class (if applicable): III		

**OIML Certificate No.
R76/2006-A-DK2-26.07**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type examination report: No. 125-29642.10, dated 10 June 2026, that includes 68 pages

Type Evaluation report: No. 125-29642.90.10, dated 04 August 2026, 22 pages.

The technical documentation relating to the identified type is contained in documentation file:
125-29642

OIML Certificate History

Revision No.	Date	Description of the modification
Initial version	04 August 2026	-

Identification, signature and stamp
The OIML Issuing Authority
FORCE Certification A/S

Date: 04 August 2026

Michael Lang Sørensen
Certification Manager

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Descriptive annex

Characteristics

Type:	X One Series
Accuracy Class:	III
Weighing ranges:	Single-interval or multi-interval
Maximum capacity (Max):	6 kg to 30 kg
Minimum capacity (Min):	20 e
Maximum number of verification scale intervals(n):	6000 (single interval) or 2×3000 (multi-interval)
Verification scale interval (e=):	≥ 1 g
Maximum tare effect:	≤ -Max
Load cell:	Avery T103
Operating temperature range:	-10 °C / +40 °C
Power supply:	110 – 240 VAC 50/60 Hz

Software

The software versions are identified with the part number and version of each part separately.

Application Software	ABR30-000495
Application	v6.x.x.x
Validation Library	v2.x.x.x
Legally Relevant Library	v3.x.x.x
Load Cell Software	ABR30-000497 v6.x.x.x
Printer Software	ABR30-000497 v6.x.x.x

Devices

- Initial zero setting device ($\leq 20\%$ of Max)
- Semi-automatic zero setting device ($\leq 4\%$ of Max)
- Automatic zero setting device ($\leq 4\%$ of Max)
- Zero tracking device ($\leq 4\%$ of Max)
- Semi-automatic subtractive tare device, balancing or weighing
- Preset tare device
- Price-computing device
- Printing device
- Price-labelling device
- Levelling device
- Gravity compensation device (sealed)

Interfaces

Communication interfaces:

- Ethernet
- Wi-Fi
- USB

The interface does not have to be secured.

