



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.42 revision 1
Project number 3969424
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Ohaus Corporation
8 Campus Drive, Suite 105
Parsippany New Jersey, 07054
United States of America

Identification of the
certified type

A Non-automatic weighing instrument
Type

: Scout SJX series
(SJX323M, SJX622M, SJX1502M,
SJX6201M)

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.

This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
1 June 2026

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.42 revision 1
Project number 3969424
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-15200577-01 dated 1 September 2016 that includes 44 pages;
- No. NMI-15200577-02 dated 1 September 2016 that includes 16 pages;
- No. NMI-15200577-03 dated 1 September 2016 that includes 32 pages;
- No. NMI-15200577-04 dated 1 September 2016 that includes 24 pages;
- No. NMI-1901086-01 dated 4 July 2017 that includes 20 pages;
- No. NMI-1902590-01 dated 2 October 2018 that includes 10 pages;
- No. NMI-3969424-01 dated 1 June 2026 that includes 18 pages;
- No. NMI-3969424-02 dated 1 June 2026 that includes 13 pages.

Characteristics of the non-automatic weighing instrument:

Type	SJX323M	SJX622M	SJX1502	SJX6102M
Accuracy class	II			
Maximum capacity	Max ≤ 64 g, or Max ≤ 320 ct	Max ≤ 620 g, or Max ≤ 3100 ct	Max ≤ 1500 g, or Max ≤ 7500 ct	Max ≤ 6200 g, or Max ≤ 31000 ct
Verification scale interval	e = 0,01 g, or e = 0,1 ct	e = 0,1 g, or e = 0,5 ct	e = 0,1 g, or e = 0,5 ct	e = 1 g, or e = 5 ct
Actual scale interval	e = d, or e = 10 d			
Weighing ranges	Single interval			
Maximum number of scale intervals	n ≤ 6400	n ≤ 6200	n ≤ 15000	n ≤ 6200
Tare	T ≤ -Max			
Temperature range	+10 °C / +30 °C			
Power supply voltage	6 V DC (battery), or 5 V DC (supplied by AC/DC adapter, or USB)			

Software identification	Mainboard only		Sr 1.xx (xx is 00 to 99 and covers the non legally relevant software)
	Mainboard 31065529 with cell board	Cell board	LC 1.x.x (x is 0 to 99 and covers the non legally relevant software)
		Mainboard	Sr 2.xx (xx is 00 to 99 and covers the non legally relevant software)



OIML Member State
The Netherlands

OIML Certificate

Number R76/2006-A-NL1-18.42 revision 1
Project number 3969424
Page 3 of 3

Revision History

+ This revision replaces the previous version.

Revision	Date	Change(s)
0	2018-10-02	Initial issue of the OIML Certificate under scheme A
1	2026-06-01	Added the reports for the addition of main board 31065529 and cell board including a new A/D chip for temperature compensation. Updated the software identification.