



# OIML Certificate

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-20.07 revision 1  
Project number 2683156  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

Cascade Corporation  
2201 NE 201<sup>st</sup> Ave  
Fairview, OR 97024  
USA

Identification of the  
certified type

**A Non-automatic weighing instrument**  
Type

: 23C-WFx-2A-yyyyy  
28C-WFx-3A-yyyyy  
50C-WFx-3A-yyyyy  
where x = E, R or G, y = 0-9R

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **III** or **IIII**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
23 July 2025

Certification Board

NMi Certin B.V.  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.





# OIML Certificate

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-20.07 revision 1  
Project number 2683156  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-2683156-07 dated 23 July 2025 that includes 8 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Maximum capacity	$1000 \text{ kg} \leq \text{Max} \leq 4500 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ kg}$
Weighing ranges	Single interval Multi-interval
Maximum number of scale intervals	$n \leq 2500$
Maximum number of partial weighing ranges	3
Tare	$T \leq -\text{Max} \%$ for instruments with one weighing range

## Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
0	2020-02-19	Initial issue.
1	2025-07-23	Correcting Software versions and procedural improvements.