



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-20.50 revision 1
Project number 3779824
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Dibal S.A.
Astintze 26
48160 Derio
Spain

Identification of the
certified type

A Non-automatic weighing instrument
Type : CS Series

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
13 February 2024

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-20.50 revision 1
Project number 3779824
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-1902587-01 dated 14 December 2018 that includes 53 pages;
- No. NMI-1902587-02 dated 14 December 2018 that includes 12 pages;
- No. NMI-1902587-03 dated 14 December 2018 that includes 12 pages;
- No. NMI-2431310-01 dated 18 March 2020 that includes 12 pages;
- No. NMI-2532570-01 dated 30 December 2020 that includes 13 pages;
- No. NMI-3700931-01 dated 15 November 2023 that includes 12 pages;
- No. NMI-3700931-02 dated 15 November 2023 that includes 22 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		III
Maximum capacity		6 kg ≤ Max ≤ 100 kg
Verification scale interval		e ≥ 2 g
Weighing ranges		Single interval Multi-interval Multiple range
Maximum number of scale intervals (one weighing range)		n ≤ 6000
Maximum number of scale intervals (multi-interval)		n ≤ 3000 (per partial weighing range)
Maximum number of partial weighing ranges		2
Maximum number of scale intervals (multiple range)		n ≤ 3000 (per weighing range)
Maximum number of weighing ranges		2
Tare		Tare ≤ -Max + e for single interval Tare ≤ -Max ₁ + e ₁ for multi-interval and multiple range instruments
Temperature range		0 °C / 40 °C
Power supply voltage		110 – 230 V AC 50/60 Hz
Application		Intended to be used for direct sales to the public
Software identification	Weighing CPU Firmware	Version : 2.0.3
	PC Software checksums	PHP : F244B8F0
		JPOS : ADD09326
		Net: 2ADDCC7C



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-20.50 revision 1
Project number 3779824
Page 3 of 3

Revision History

+ This revision replaces the previous version.

Revision	Date	Changes
0	2020-12-30	Initial issue.
1	2024-02-13	Addition of reports.