OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-20.62 revision 1
Project number 4158817

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Schmidt
Applicant and Manufacturer	Dibal S.A. Astintze Kalea 26 48160 Derio Spain
Identification of the certified type	A **Non-automatic weighing instrument** Type : KS-400 / KS-200 / KS Series
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
5 October 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.

OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-20.62 revision 1
Project number 4158817

The conformity was established by the results of tests and examinations provided in the associated reports:
- No. NMi-16200563-01 dated 14 March 2017 that includes 60 pages;
- No. NMi-16200563-02 dated 14 March 2017 that includes 32 pages.
- No. NMi-2532564-01 dated 4 February 2021 that includes 20 pages.

Characteristics of the non-automatic weighing instrument:

The construction of the non-automatic weighing instrument is recorded in the Documentation folder no. T10856-3.

Accuracy class	(III)
Maximum capacity	6 kg ≤ Max ≤ 30 kg
Verification scale interval	e ≥ 2 g
Weighing ranges	Single interval Multiple range
Maximum number of partial weighing ranges	2
Maximum number of scale intervals (single interval / multiple range)	n ≤ 3000 divisions (per weighing range)
Maximum number of weighing ranges	2
Tare	T ≤ Max - e for single interval instruments T ≤ Max_1 - e_1 for multiple range instruments
Temperature range	+ 5 °C / 40 °C
Power supply voltage	5-12 V DC via plug in power supply 5 V via USB 100 – 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public
Software identification	Legally relevant software version number: V-1.01

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
0	2021-02-21	Initial issue.
1	2026-10-05	Update type name