



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.10 revision 1
Project number 3914466
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

InBody Co., Ltd.
15, Heugam-gil, Ipjang-myeon, Seobuk-gu
Cheonan-si
Chungcheongnam-do 31025
Republic of Korea

Identification of the
certified type

A Non-automatic weighing instrument
Type : InBody970 / InBody970S

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
24 October 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.





OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.10 revision 1
Project number 3914466
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-2446349-01 dated 26 February 2021 that includes 54 pages;
- No. NMI-3914466-01 dated 24 October 2025 that includes 16 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	300 kg
Verification scale interval	e = 100 g
Weighing range	Single interval
Maximum number of scale intervals	n = 3000
Temperature range	+10 °C / +40 °C
Power supply voltage	12 V DC from 100-240 V AC 50/60 Hz plug-in power supply
Application	Intended to be used for weighing patients

Type	InBody970	InBody970S
Software identification	IB970-1.x (x = 0... 9 and represents the non-legally relevant software)	IB970S-1.x (x = 0... 9 and represents the non-legally relevant software)
	970DM-0421	970SDM-0421
Load Cell Board Firmware version		970LCB-C013

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
0	2021-02-21	Initial issue.
1	2025-10-24	Added InBody970S, added supplementary legally relevant software versions. Added the report NMI-3914466-01 for the evaluation of the supplementary legally relevant software modules.