



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R76/2006-A-NL1-21.24 revision 3  
Project number 3741110  
Page 1 of 3

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

Bizerba SE & Co. KG  
Wilhelm-Kraut-Str. 65  
72336 Balingen  
Germany

Identification of the  
certified type

**A Non-automatic weighing instrument**

Type : Q1...100... (bench type)  
Q1...100... (pole type)  
Q1...300... (bench type)  
Q1...400... (hanging type)  
Q1...800... (pole type)  
'...' denotes a placeholder for  
alternative variants not affecting  
metrological parameters.

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
12 March 2024

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.





**OIML Member State**  
The Netherlands

# OIML Certificate

Number R76/2006-A-NL1-21.24 revision 3  
Project number 3741110  
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

- No. NMI-2477454-01 dated 15 April 2021 that includes 51 pages;
- No. NMI-2477454-02 dated 15 April 2021 that includes 18 pages;
- No. NMI-2529842-01 dated 15 April 2021 that includes 17 pages;
- No. NMI-3001387-01 dated 18 October 2022 that includes 26 pages;
- No. NMI-3001387-02 dated 18 October 2022 that includes 14 pages;
- No. NMI-3741110-01 dated 12 March 2024 that includes 22 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class		III
Maximum capacity	Weighing module QBase	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
	Weighing module ScalePad USB	$\text{Max} \leq 15 \text{ kg}$
	Weighing module ADC-Box USB	Calculated using the compatibility of modules form, contained in EN 45501:2015 clause F.4
Verification scale interval		$e \geq 1 \text{ g}$
Weighing ranges		Single interval Multi-interval
Maximum number of scale intervals		$n \leq 6000$
Maximum number of scale intervals (multi-interval)		$n \leq 3000$ (per partial weighing range)
Maximum number of partial weighing ranges		2
Semi-automatic Tare		$T \leq -100\% \text{ of Max}$
Preset Tare		100% of Max; 100% of $\text{Max}_1$ , in case of multi-interval instruments
Temperature range		$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage		120 – 240 V AC 50/60 Hz
Application		Intended to be used for direct sales to the public

## Terminal software identification:

Scale OEM Module	Software version number (legally relevant) <sup>1)</sup>	Software ID (Windows)	Software ID (Linux)
	013	7640	0369
	015	8282	1171

## A/D board software identification:

ADW	Software version number (legally relevant) <sup>1)</sup>	Software ID
	010	6153
	011	5232
	030	1446



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R76/2006-A-NL1-21.24 revision 3  
Project number 3741110  
Page 3 of 3

- <sup>1)</sup> The identification of the non-legally relevant software version can optionally be added to the identification of the legally relevant software version, separated by "::" (example: "010::x.y" or "015::x.y").

The software identifications will be shown after tapping the part of the display with the markings Max, Min and e for at least 5 seconds.

## Revision History

This revision replaces the previous versions.

Revision	Date	Change(s)
0	2021-04-15	Initial issue.
1	2022-10-18	Adding Q1...400... (hanging type) and Q1...100.... (pole type). Adding ScalePad USB and ADC-Box USB.
2	2023-01-05	Addition of a new software ID for A/D board.
3	2024-03-12	Adding Q1...300 (bench type), adding alternative power supply and I/O board