



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.24 revision 5
Project number 4000660
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Bizerba SE & Co. KG
Wilhelm-Kraut-Str. 65
72336 Balingen
Germany

Identification of the
certified type

A Non-automatic weighing instrument

Type : Q1...100... (bench type)
Q1...100... (pole type)
Q1...300... (bench type)
Q1...400... (hanging type)
Q1...800... (pole type)
'...' denotes a placeholder for alternative variants not affecting
metrological parameters.

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
18 November 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.





OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.24 revision 5
Project number 4000660
Page 2 of 4

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-2477454-01 revision 1 dated 15 September 2023 that includes 51 pages;
- No. NMI-2477454-02 dated 15 April 2021 that includes 18 pages;
- No. NMI-2529842-01 dated 15 April 2021 that includes 17 pages;
- No. NMI-3001387-01 dated 18 October 2022 that includes 26 pages;
- No. NMI-3001387-02 dated 18 October 2022 that includes 14 pages;
- No. NMI-3741110-01 dated 12 March 2024 that includes 22 pages;
- No. NMI-3741825-01 dated 29 August 2024 that includes 15 pages;
- No. NMI-4000660-01 dated 18 November 2025 that includes 20 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		III
Maximum capacity	Weighing module QBase	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
	Weighing module ScalePad USB	$\text{Max} \leq 15 \text{ kg}$
	Weighing module ADC-Box USB	Calculated using the compatibility of modules form, contained in EN 45501:2015 clause F.4
Verification scale interval		$e \geq 1 \text{ g}$
Weighing ranges		Single interval Multi-interval
Maximum number of scale intervals		$n \leq 6000$
Maximum number of scale intervals (multi-interval)		$n \leq 3000$ (per partial weighing range)
Maximum number of partial weighing ranges		2
Semi-automatic Tare		$T \leq -100\% \text{ of Max}$
Preset Tare		100% of Max; 100% of Max_1 , in case of multi-interval instruments
Temperature range		$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage		120 – 240 V AC 50/60 Hz
Application		Intended to be used for direct sales to the public

Terminal software identification:

Scale OEM Module	Software version number (legally relevant) ¹⁾	Software ID (Windows)	Software ID (Linux)
	013	7640	0369
	015	8282	1171
	050	5275	8164
	051	8010	4115



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.24 revision 5
Project number 4000660
Page 3 of 3

	060	5593	1698
--	-----	------	------

A/D board software identification:

ADW	Software version number (legally relevant) ¹⁾	Software ID
	010	6153
	011	5232
	030	1446
	014	4751

- ¹⁾ The identification of the non-legally relevant software version can optionally be added to the identification of the legally relevant software version, separated by "::
(example: "010::x.y" or "015::x.y").

The software identifications will be shown after tapping the part of the display with the markings Max, Min and e for at least 5 seconds.

Revision History

This revision replaces the previous versions.

Revision	Date	Change(s)
0	2021-04-15	Initial issue
1	2022-10-18	Adding Q1...400... (hanging type) and Q1...100.... (pole type). Adding ScalePad USB and ADC-Box USB.
2	2023-01-05	Addition of a new version.
3	2023-09-22	Replace type evaluation report NMI-2477454-01 with NMI-2477454-01 revision 1.
4	2024-01-02	Adding Q1...300 (bench type), alternative power supply, alternative I/O board and alternative A/D board with alternative software version number and software ID.
5	2025-11-18	Adding software versions and test report NMI-4000660-01