



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.45 revision 1
Project number 3749295
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

METTLER-TOLEDO Changzhou Measurement Technology Ltd.
No.111, West TaiHu Road,
Changzhou, Jiangsu, 213125
China

Identification of the
certified type

An **Indicator** or **Terminal**
Type : IND500x

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (II) or (III) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
15 February 2024

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.45 revision 1
Project number 3749295
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-2657878-01 dated 24 December 2021 that includes 50 pages;
- No. NMI-2657878-02 dated 24 December 2021 that includes 13 pages.

Characteristics of the indicator / terminal:

		Analog load cells	Digital load cells
Accuracy class	OIML R 76	Ⓐ or Ⓑ	Ⓐ or Ⓑ or Ⓒ
Weighing ranges		Single interval Multi-interval Multiple range	
Maximum number of scale intervals		$n \leq 10000$ divisions	$n \leq 100000$ divisions
Load cell excitation voltage		4,8 V DC	-
Load cell power supply		-	8,2 / 12 V DC
Minimum input voltage per verification scale interval		0,3 μ V	-
Minimum load cell resistance		80 Ω	-
Maximum load cell resistance		3000 Ω	-
Fraction of the maximum permissible error		0,5	0
Load cell interface		6-wire with sense technology, may be configured as 4-wire	-
Maximum value of the cable length per cross wire section between the indicator and the junction box or load cells		4487 m/mm ²	-
Maximum number of load platforms		1	
Climatic environment	temperature range	-10 °C / +40 °C	
	humidity	non-condensing	
	intended location	Closed	
Electromagnetic environment class		E2	
Power supply voltage		5,6 V, 8,2 V and 12,6 V DC through AC/DC power supply	



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.45 revision 1
Project number 3749295
Page 3 of 3

Software identification:

Description	Version	Remarks
Standard	1.xx.yyyy	For analog, IDNet and SICSpro scale types
Standard	2.xx.yyyy	Source code of LFT relevant files is optimized & modified but without changing the behaviour of legally relevant functions

(xx is a number between 00 and 99 representing major updates of the non-legally relevant part of the software and yyyy is a number between 0000 and 9999 and represents minor updates of the non-legally relevant part of the software)

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
0	2021-12-24	Initial issue.
1	2024-02-15	Adding software version 2.xx.yyyy