



# OIML Certificate

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-23.04 revision 1  
Project number 3760795  
Page 1 of 3

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

Zebra Technologies Corp  
1 Zebra Plaza  
Holtsville, NY 11742  
United States of America

Identification of the  
certified type

**A Non-automatic weighing instrument**

Type : MP70X1 single interval  
MP70X2 multi-interval  
MP72X1 single interval  
MP72X2 multi-interval  
MP72X3 single-interval with calibration push button  
MP72X4 multi-interval with calibration push button

'X' denotes alternative variants not affecting  
metrological parameters.

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
26 April 2024

Certification Board

NMi Certin B.V.  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.





# OIML Certificate

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-23.04 revision 1  
Project number 3760795  
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. SN 1379 dated 12 May 2017 that includes 47 pages;
- No. P02070 dated 12 May 2017 that includes 14 pages;
- No. P02848 dated 8 October 2020 that includes 16 pages;
- No. NMI-2752431-01 dated 17 February 2023 that includes 45 pages;
- No. NMI-2752431-02 dated 17 February 2023 that includes 20 pages;
- No. NMI-3760795-01 dated 26 April 2024 that includes 21 pages.

## Characteristics of the non-automatic weighing instrument:

|   |  |
|---|--|
| Accuracy class  | III  |
| Maximum capacity  | Max $\leq 15$ kg                                   |
| Verification scale interval                               | $e \geq 2$ g                                       |
| Weighing ranges   | Single interval<br>Multi-interval                  |
| Maximum number of scale intervals<br>(one weighing range) | $n \leq 3000$                                      |
| Maximum number of scale intervals<br>(multi-interval)     | $n \leq 3000$<br>(per partial weighing range)      |
| Maximum number of<br>partial weighing ranges              | 2  |
| Temperature range   | 0 °C / +40 °C                                      |
| Power supply voltage                                      | 100 – 240 V AC 50/60 Hz<br>12 V DC                 |
| Application   | Intended to be used for direct sales to the public |
| Software<br>identification                                | Version number<br>1.04F                            |

When pressing and holding the zero button (<0>) for a period of at least 3 seconds at stable condition, the operator display will display:

- Software identification;
- Calibration counter (in the form "C xxx");
- Parameter counter (in the form "P xxx").



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R76/2006-A-NL1-23.04 revision 1  
Project number 3760795  
Page 3 of 3

## Revision History

This revision replaces the previous version.

| Revision | Date       | Change(s)  |
|----------|------------|--|
| 0        | 2023-02-17 | Initial issue.   |
| 1        | 2024-04-26 | Adding new model MP72 and alternative main board PA-00204-XX |