



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-23.11 revision 3
Project number 4000660
Page 1 of 4

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Bizerba SE & Co. KG
Wilhelm-Kraut-Str. 65
72336 Balingen
Germany

Identification of the
certified type

A Non-automatic weighing instrument
Type : C2 series

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
18 November 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.





OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-23.11 revision 3
Project number 4000660
Page 2 of 4

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-2431931-01 dated 14 April 2020 that includes 45 pages.
- No. NMI-3532583-01 dated 14 April 2023 that includes 15 pages;
- No. NMI-3532583-02 dated 14 April 2023 that includes 21 pages;
- No. NMI-3532583-03 dated 14 April 2023 that includes 21 pages;
- No. NMI-3680753-01 dated 21 December 2023 that includes 36 pages;
- No. NMI-3680753-02 dated 21 December 2023 that includes 17 pages;
- No. NMI-3741825-01 dated 29 August 2024 that includes 15 pages;
- No. NMI-4000660-01 dated 18 November 2025 that includes 20 pages.

Characteristics of the non-automatic weighing instrument:

| | | |
|--|---------------------------------|--|
| Accuracy class | | III |
| Weighing range | | Single interval Multi-interval |
| Maximum capacity | Weighing module QBase | $\text{Max} \leq 30 \text{ kg}$ |
| | Weighing module scanner variant | $\text{Max} \leq 15 \text{ kg}$ |
| Verification scale interval | Weighing module QBase | $e \geq 1 \text{ g}$ |
| | Weighing module scanner variant | $e \geq 2 \text{ g}$ |
| Maximum number of scale intervals (one weighing range) | | $n \leq 6000$ |
| Maximum number of scale intervals (multi-interval) | | $n \leq 3000$ (per partial weighing range) |
| Maximum number of weighing ranges | | 2 |
| Tare | | $T \leq - \text{Max}$ for single interval instruments and $\text{Max} < 10 \text{ kg}$ |
| | | $T \leq - (10 \text{ kg} - e)$ for single interval instruments and $\text{Max} \geq 10 \text{ kg}$ |
| | | $T \leq - \text{Max}_2$ for multi-interval instrument and $\text{Max}_2 < 10 \text{ kg}$ |
| | | $T \leq - (10 \text{ kg} - e_2)$ for multi-interval instrument and $\text{Max}_2 \geq 10 \text{ kg}$ |
| Temperature range | | $-10 \text{ }^{\circ}\text{C} / +40 \text{ }^{\circ}\text{C}$ |
| Power supply voltage | | 12 V DC |
| Application | | Intended to be used for direct sales to the public |

OIML Member State
The Netherlands

Number R76/2006-A-NL1-23.11 revision 3
Project number 4000660
Page 3 of 4

Terminal software identification:

| | Software version number (legally relevant) ¹⁾ | Software ID (Windows) | Software ID (Linux) |
|------------------|---|-----------------------|---------------------|
| Scale OEM Module | 013 | 7640 | 0369 |
| | 015 | 8282 | 1171 |
| | 050 | 5275 | 8164 |
| | 051 | 8010 | 4115 |
| | 060 | 5593 | 1698 |

Weighing module software identification:

| Scale software | Software version number (legally relevant) ¹⁾ | Software ID |
|----------------|--|-------------|
| ADW | 010 | 6153 |
| | 011 | 5232 |
| | 030 | 1446 |
| | 014 | 4751 |
| CPU | 001 | 8324 |
| | 002 | 5242 |

¹⁾ The identification of the non-legally relevant software version can optionally be added to the identification of the legally relevant software version, separated by "::" (example: "011::x.y" or "010::x.y" or "030::x.y" or "014::x.y").

Software identification:

The software version number and software ID can be accessed on the Scale OEM module by pressing the inscriptions "Min, Max and e" for at least 3 seconds.

Additionally the software version number and software ID can be displayed using the scanner buttons (depending on terminal software version):

- Press keys "0, 0, 0" and then "T" key to open the logbook;
- Use "T" key to navigate the logbook. The key "T" can be different symbols (e.g. "volume" or "!") depending on the actual scanner model. For detailed instructions check the manual of the weighing instrument.
- Press key "0" to close the logbook.

In case a C2 terminal is used, the software version number and software ID are displayed after pressing the following key sequence with the "0" and "T" keys:

- press keys "0, 0, 0, T" during start-up procedure of the instrument;
 - Use key "0" to scroll to "Info" and confirm with key "T";
 - Use key "0" to scroll to "C2" or "AdC", use key "T" to select and enter log data;
 - Use key "0" to scroll through the log data, use key "T" to step back from detailed view;
 - The software version number is shown at "c2,Lv" (CPU board) or "AdC,Lv" (A/D board);
- The software ID is shown at "c2,i" (CPU board) or "AdC,i" (A/D board).



OIML Member State
The Netherlands

OIML Certificate

Number R76/2006-A-NL1-23.11 revision 3
Project number 4000660
Page 4 of 4

Revision History

This revision replaces the previous versions.

| Revision | Date | Changes |
|----------|------------|---|
| 0 | 2023-04-14 | Initial issue. |
| 1 | 2023-12-21 | Change in the characteristics table and adding test reports. |
| 2 | 2025-01-02 | Addition of an alternative A/D board and terminal software version numbers. |
| 3 | 2025-11-18 | Adding software versions and test report NMI-4000660-01 |