



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-24.01 revision 0
Project number 3755260
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Fook Tin Technologies Ltd.
4/F., Eastern Ctr.
1065 King's Rd, Quarry Bay
Hong Kong

Identification of the
certified type

A Non-automatic weighing instrument
Type : FT-AERO Baggage scale / Bag2R
Baggage scale / FP-BS.... Series

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
5 February 2024

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.





OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-24.01 revision 0
Project number 3755260
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

Indicator:

AWI-P / AWI-K / FT-AERO-1 / FT-AERO-2 / ASD-1 / ASD-2

- No. 120-32716.10 dated 18 January 2021 that includes 65 pages;
- No. 120-32716.90.20 dated 8 February 2021 that includes 22 pages.

Analog load cell:

SK30X...

- No. R60/2000-NL1-04.04 dated 11 March 2004 that includes 40 pages;
- No. R60/2000-NL1-05.10 dated 12 August 2005 that includes 37 pages;
- No. R60/2000-NL1-06.05 dated 16 May 2006 that includes 37 pages;
- No. R60/2000-NL1-06.13 dated 29 August 2006 that includes 37 pages.

Digital load cell:

SDX

- No. NMI-3579824-01 dated 15 June 2023 that includes 61 pages;
- No. NMI-3579824-02 dated 15 June 2023 that includes 15 pages.

For each model, the compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in OIML R76 edition 2006 Annex F.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III	
Maximum capacity	Max \leq 300 kg	
Verification scale interval	$e \geq 0,1$ kg	
Weighing range	Single interval	
Maximum number of scale intervals	$n \leq 3000$	
Maximum number of weighing ranges	1	
Tare	$T \leq -50\%$	
Temperature range	$-10\text{ }^{\circ}\text{C} / 40\text{ }^{\circ}\text{C}$	
Power supply voltage	12 V DC, 110 – 230 V AC 50/60 Hz	
Software identification	Version number	v1.0 (for indicator) v2.7 (for digital load cell)
	Checksum	41EC (for indicator) 0xAE57 (for digital load cell)



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-24.01 revision 0
Project number 3755260
Page 3 of 3

Revision History

Revision	Date	Change
0	5 February 2024	Initial issue.