

OIML Member State

The Netherlands

OIML Certificate



Number R76/2006-A-NL1-24.06 revision 0 Project number 3660181 Page 1 of 2

Issuing authority NMi Certin B.V. Person responsible: M.Ph.D. Schmidt Applicant and InBody Co., Ltd. Manufacturer 15, Heugam-gil, Ipjang-myeon, Seobuk-gu Cheonan-si Chungcheongnam-do 31025 Korea A Non-automatic weighing instrument Identification of the certified type : InBody 380, InBody 580 Type Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority



NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl NMi Certin B.V., OIML Issuing Authority NL1 4 April 2024

Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.







OIML Certificate

OIML Member State The Netherlands



Number R76/2006-A-NL1-24.06 revision 0 Project number 3660181 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMi-2446349-01 dated 26 February 2021 that includes 54 pages;
 - No. NMi-3660181-01 dated 4 April 2024 that includes 19 pages;
 - No. NMi-3660181-02 dated 4 April 2024 that includes 21 pages;
 - No. NMi-3660181-03 dated 4 April 2024 that includes 15 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		
Maximum capacity		300 kg
Verification scale interval		e ≥ 100 g
Weighing range(s)		Single interval
Maximum number of scale intervals (one weighing range)		n ≤ 3000
Temperature range		+10 °C / +40 °C
Power supply voltage		100 – 240 V AC 50/60 Hz
Application		Intended to be used for weighing patients
Software	Main software version	IB380-1.0
identification Inbody 380	Data management software version	380DM-0421
Software identification Inbody 580	Main software version	IB580-1.0
	Data management software version	580DM-0421
Load Cell Board Firmware version		970LCB-C013

Revision History

Revision	Date	Change
0	2024-04-04	Initial issue.