



# OIML Certificate

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-24.30 revision 1  
Project number 3922182  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

Shanghai Ishida Electronic Scales Co., Ltd.  
Building 2, No.86, MinXue Road  
Pudong Shanghai  
China

Identification of the  
certified type

**A Non-automatic weighing instrument**  
Type : IPC / s-boxR

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **III** or **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
30 January 2025

Certification Board

NMi Certin B.V.  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-24.30 revision 1  
Project number 3922182  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3806248-01 dated 6 January 2025 that includes 47 pages;
- No. NMI-3806248-02 dated 6 January 2025 that includes 19 pages;
- No. NMI-3806248-03 dated 6 January 2025 that includes 18 pages;
- No. NMI-3922182-01 dated 30 January 2025 that includes 9 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Maximum capacity	$1,5 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	$n \leq 1500$
Maximum number of scale intervals (multi-interval)	$n \leq 1500$ (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	$T \leq -50\%$ for instruments with one weighing range $T \leq -\text{Max}_1 - e_1$ for multi-interval instruments
Temperature range	$-5^\circ\text{C} / +40^\circ\text{C}$
Power supply voltage	230 V AC 50/60 Hz 3 V DC by battery
Software identification      Version number	L004

The software identification is displayed at start-up.

The non-automatic weighing instrument has embedded software.

## Revision History

Revision	Date	Change(s)
0	2025-01-06	Initial issue.
1	2025-01-30	Addition of type evaluation report.