



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.05 revision 1
Project number 4160734
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

DIGI SINGAPORE PTE LTD.
4 Leng Kee Road, #06-01
SIS Building, 159088
Singapore

Identification of the
certified type

A Non-automatic weighing instrument
Type : DRS-1000, DRS-1000E

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
13 August 2026

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
the Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.





OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.05 revision 1
Project number 4160734
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-2733270-01 dated 25 February 2025 that includes 51 pages;
- No. NMI-3565144-01 dated 16 June 2023 that includes 27 pages;
- No. NMI-3793517-01 dated 15 April 2025 that includes 15 pages;
- No. NMI-3793517-02 dated 15 April 2025 that includes 16 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III	
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$	
Verification scale interval	$e \geq 1 \text{ g}$	
Weighing range	Single interval	
Maximum number of scale intervals	$n \leq 6000$	
Temperature range	$- 10 \text{ }^{\circ}\text{C} / + 40 \text{ }^{\circ}\text{C}$	
Power supply voltage	12 – 24 V DC supplied over the Ethernet interface	
Application	Intended to be used for direct sales to the public	
Software identification	Version number	V1.xx or V2.xx (xx = 00... 99 and represents the non-legally relevant part of the software)

The software identification is displayed at start-up.

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
0	2025-04-15	Initial issue.
1	2026-08-13	Addition of type name DRS-1000E.