



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.08 revision 0
Project number 3937345
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

SAT America Inc.
7226 Nw 56th St Miami
33166 Miami, Florida
United States of America

Identification of the
certified type

A Non-automatic weighing instrument
Type : SAT PS30CExx
Brand : SAT PCS

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
3 June 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.08 revision 0
Project number 3937345
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-10200947-07 dated 24 December 2010 that includes 59 pages;
- No. NMI-10200947-08 dated 24 December 2010 that includes 59 pages;
- No. R60/2000-NL1-10-07 dated 17 May 2010 that includes 40 pages;
- No. NMI-12200582-01 dated 23 August 2013 that includes 52 pages;
- No. NMI-10200906-01 dated 20 April 2011 that includes 36 pages;
- No. NMI-11200490-01 dated 27 September 2011 that includes 39 pages;
- No. NMI-SO11201010-01 dated 30 November 2011 that includes 9 pages;
- No. NMI-12200120-01 dated 13 September 2012 that includes 13 pages;
- No. NMI-12200120-02 dated 13 September 2012 that includes 12 pages;
- No. NMI-12200787-01 dated 15 March 2013 that includes 9 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 2 \text{ g}$
Weighing ranges	Single interval Multi-interval Multiple range
Maximum number of scale intervals	$n \leq 3000$ divisions (per partial weighing range)
Maximum partial weighing ranges	2
Temperature range	$0 \text{ }^{\circ}\text{C} / +40 \text{ }^{\circ}\text{C}$
Tare	$T \leq -(\text{Max} - e)$ for single interval instruments $T \leq -(\text{Max}_1 - e_1)$ for multi-interval instruments
Power supply voltage	110 – 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public and for the making-up of prepackages
Software identification	Version number: V1.045, or V1.049, or V3.072

Revision History

Revision	Date	Change(s)
0	2025-06-03	Initial issue.